



ANALYSIS REPORT

Prepared by:

Eurofins Lancaster Laboratories Environmental
2425 New Holland Pike
Lancaster, PA 17601

Prepared for:

C. T. Male Associates
50 Century Hill Drive
Latham NY 12110

Report Date: December 30, 2019 08:44

Project: Hoosick Falls WTP

Account #: 37191
Group Number: 2078432
SDG: HOO40
PO Number: 14.4756
State of Sample Origin: NY

Electronic Copy To	C. T. Male Associates	Attn: Kirk Moline
Electronic Copy To	C. T. Male Associates	Attn: Dan Reilly
Electronic Copy To	C. T. Male Associates	Attn: Jeff Marx
Electronic Copy To	Barr Engineering Company	Attn: Lauren Brady
Electronic Copy To	Environmental Standards	Attn: St. Gobain
Electronic Copy To	Barr Engineering Company	Attn: Data Mgt
Electronic Copy To	Barr Engineering Company	Attn: Terri Olson
Electronic Copy To	C. T. Male Associates	Attn: Nancy Garry

Respectfully Submitted,



(717)-556-7376

To view our laboratory's current scopes of accreditation please go to <https://www.eurofinsus.com/environment-testing/laboratories/eurofins-lancaster-laboratories-environmental/certifications-and-accreditations-eurofins-lancaster-laboratories-environmental/> . Historical copies may be requested through your project manager.



SAMPLE INFORMATION

<u>Client Sample Description</u>	<u>Sample Collection Date/Time</u>	<u>ELLE#</u>
LTB01-191209 Blank Water	12/09/2019	1219261
GAC Influent Grab Drinking Water	12/09/2019 09:55	1219262
GAC Midfluent Grab Drinking Water	12/09/2019 09:57	1219263
GAC Effluent Grab Drinking Water	12/09/2019 10:00	1219264
PV-1 50 Grab Drinking Water	12/09/2019 10:03	1219265
PV-2 50 Grab Drinking Water	12/09/2019 10:05	1219266
FTB01-191209 Grab Blank Water	12/09/2019 10:10	1219267

The specific methodologies used in obtaining the enclosed analytical results are indicated on the Laboratory Sample Analysis Record.

Project Name: Hoosick Falls WTP
ELLE Group #: 2078432

General Comments:

See the Laboratory Sample Analysis Record section of the Analysis Report for the method references.

All QC met criteria unless otherwise noted in an Analysis Specific Comment below.

Refer to the QC Summary for specific values and acceptance criteria.

Project specific QC samples are not included in this data set.

Matrix QC may not be reported if site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD was performed, unless otherwise specified in the method.

Surrogate recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in an Analysis Specific Comment below.

The samples were received at the appropriate temperature and in accordance with the chain of custody unless otherwise noted.

Analysis Specific Comments:

No additional comments are necessary.

Sample Description: LTB01-191209 Blank Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: WW 1219261
ELLE Group #: 2078432
Matrix: Blank Water

Project Name: Hoosick Falls WTP

Submittal Date/Time: 12/10/2019 10:19
Collection Date/Time: 12/09/2019
SDG#: HOO40-01TB

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.7 U	1.7	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.7 U	1.7	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.7 U	1.7	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.7 U	1.7	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.7 U	1.7	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.7 U	1.7	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.7 U	1.7	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.7 U	1.7	1
14070	Perfluorononanoic acid ¹	375-95-1	1.7 U	1.7	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.7 U	1.7	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.7 U	1.7	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.7 U	1.7	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.7 U	1.7	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.7 U	1.7	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.4 U	4.4	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.2 U	6.2	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	19346028	12/18/2019 05:01	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	19351014	12/19/2019 02:23	Jason W Knight	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	19346028	12/12/2019 16:00	Isaac Phillips-Cary	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	19351014	12/17/2019 18:00	Anthony C Polaski	1

Sample Description: GAC Influent Grab Drinking Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: PW 1219262
ELLE Group #: 2078432
Matrix: Drinking Water

Project Name: Hoosick Falls WTP

Submittal Date/Time: 12/10/2019 10:19
Collection Date/Time: 12/09/2019 09:55
SDG#: HOO40-02

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1					
			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.8 U	1.8	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.8 U	1.8	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.8 U	1.8	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.8 U	1.8	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.8 U	1.8	1
14070	Perfluoroheptanoic acid¹	375-85-9	7.2	1.8	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.8 U	1.8	1
14070	Perfluorohexanoic acid¹	307-24-4	8.9	1.8	1
14070	Perfluorononanoic acid ¹	375-95-1	1.8 U	1.8	1
14070	Perfluorooctanesulfonic acid¹	1763-23-1	5.0	1.8	1
14070	Perfluorooctanoic acid¹	335-67-1	250	18	10
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.8 U	1.8	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.8 U	1.8	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.8 U	1.8	1

LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified					
			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.6 U	4.6	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.8 U	2.8	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.5 U	6.5	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid¹	2706-90-3	2.8	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	19346028	12/18/2019 05:12	Marissa C Drexinger	1
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	19346028	12/18/2019 20:11	Marissa C Drexinger	10
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	19351014	12/19/2019 02:32	Jason W Knight	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	19346028	12/12/2019 16:00	Isaac Phillips-Cary	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	19351014	12/17/2019 18:00	Anthony C Polaski	1

Sample Description: GAC Midfluent Grab Drinking Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: PW 1219263
ELLE Group #: 2078432
Matrix: Drinking Water

Project Name: Hoosick Falls WTP

Submission Date/Time: 12/10/2019 10:19
Collection Date/Time: 12/09/2019 09:57
SDG#: HOO40-03

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.7 U	1.7	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.7 U	1.7	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.7 U	1.7	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.7 U	1.7	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.7 U	1.7	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.7 U	1.7	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.7 U	1.7	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.7 U	1.7	1
14070	Perfluorononanoic acid ¹	375-95-1	1.7 U	1.7	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.7 U	1.7	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.7 U	1.7	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.7 U	1.7	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.7 U	1.7	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.7 U	1.7	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.5 U	4.5	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.3 U	6.3	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	19346028	12/18/2019 05:24	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	19351014	12/19/2019 02:41	Jason W Knight	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	19346028	12/12/2019 16:00	Isaac Phillips-Cary	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	19351014	12/17/2019 18:00	Anthony C Polaski	1

Sample Description: GAC Effluent Grab Drinking Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: PW 1219264
ELLE Group #: 2078432
Matrix: Drinking Water

Project Name: Hoosick Falls WTP

Submission Date/Time: 12/10/2019 10:19
Collection Date/Time: 12/09/2019 10:00
SDG#: HOO40-04

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.7 U	1.7	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.7 U	1.7	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.7 U	1.7	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.7 U	1.7	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.7 U	1.7	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.7 U	1.7	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.7 U	1.7	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.7 U	1.7	1
14070	Perfluorononanoic acid ¹	375-95-1	1.7 U	1.7	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.7 U	1.7	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.7 U	1.7	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.7 U	1.7	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.7 U	1.7	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.7 U	1.7	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.6 U	4.6	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.4 U	6.4	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	19346028	12/18/2019 05:35	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	19351014	12/19/2019 02:50	Jason W Knight	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	19346028	12/12/2019 16:00	Isaac Phillips-Cary	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	19351014	12/17/2019 18:00	Anthony C Polaski	1

Sample Description: PV-1 50 Grab Drinking Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: PW 1219265
ELLE Group #: 2078432
Matrix: Drinking Water

Project Name: Hoosick Falls WTP

Submission Date/Time: 12/10/2019 10:19
Collection Date/Time: 12/09/2019 10:03
SDG#: HOO40-05

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.7 U	1.7	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.7 U	1.7	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.7 U	1.7	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.7 U	1.7	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.7 U	1.7	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.7 U	1.7	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.7 U	1.7	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.7 U	1.7	1
14070	Perfluorononanoic acid ¹	375-95-1	1.7 U	1.7	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.7 U	1.7	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.7 U	1.7	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.7 U	1.7	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.7 U	1.7	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.7 U	1.7	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.5 U	4.5	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.3 U	6.3	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	19346028	12/18/2019 05:58	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	19351014	12/19/2019 02:59	Jason W Knight	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	19346028	12/12/2019 16:00	Isaac Phillips-Cary	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	19351014	12/17/2019 18:00	Anthony C Polaski	1

Sample Description: PV-2 50 Grab Drinking Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: PW 1219266
ELLE Group #: 2078432
Matrix: Drinking Water

Project Name: Hoosick Falls WTP

Submission Date/Time: 12/10/2019 10:19
Collection Date/Time: 12/09/2019 10:05
SDG#: HOO40-06

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.7 U	1.7	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.7 U	1.7	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.7 U	1.7	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.7 U	1.7	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.7 U	1.7	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.7 U	1.7	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.7 U	1.7	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.7 U	1.7	1
14070	Perfluorononanoic acid ¹	375-95-1	1.7 U	1.7	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.7 U	1.7	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.7 U	1.7	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.7 U	1.7	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.7 U	1.7	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.7 U	1.7	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.6 U	4.6	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.4 U	6.4	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	19346028	12/18/2019 06:10	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	19351014	12/19/2019 03:08	Jason W Knight	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	19346028	12/12/2019 16:00	Isaac Phillips-Cary	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	19351014	12/17/2019 18:00	Anthony C Polaski	1

Sample Description: FTB01-191209 Grab Blank Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: WW 1219267
ELLE Group #: 2078432
Matrix: Blank Water

Project Name: Hoosick Falls WTP

Submission Date/Time: 12/10/2019 10:19
Collection Date/Time: 12/09/2019 10:10
SDG#: HOO40-07FB

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.8 U	1.8	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.8 U	1.8	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.8 U	1.8	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.8 U	1.8	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.8 U	1.8	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.8 U	1.8	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.8 U	1.8	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.8 U	1.8	1
14070	Perfluorononanoic acid ¹	375-95-1	1.8 U	1.8	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.8 U	1.8	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.8 U	1.8	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.8 U	1.8	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.8 U	1.8	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.8 U	1.8	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.5 U	4.5	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.3 U	6.3	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	19346028	12/18/2019 06:21	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	19351014	12/19/2019 03:26	Jason W Knight	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	19346028	12/12/2019 16:00	Isaac Phillips-Cary	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	19351014	12/17/2019 18:00	Anthony C Polaski	1

Quality Control Summary

Client Name: C. T. Male Associates
Reported: 12/30/2019 08:44

Group Number: 2078432

Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD was performed, unless otherwise specified in the method.

All Inorganic Initial Calibration and Continuing Calibration Blanks met acceptable method criteria unless otherwise noted on the Analysis Report.

Method Blank

Analysis Name	Result	LOQ
	ng/l	ng/l
Batch number: 19346028	Sample number(s): 1219261-1219267	
NEtFOSAA	2.0 U	2.0
NMeFOSAA	2.0 U	2.0
Perfluorobutanesulfonic acid	2.0 U	2.0
Perfluorodecanoic acid	2.0 U	2.0
Perfluorododecanoic acid	2.0 U	2.0
Perfluoroheptanoic acid	2.0 U	2.0
Perfluorohexanesulfonic acid	2.0 U	2.0
Perfluorohexanoic acid	2.0 U	2.0
Perfluorononanoic acid	2.0 U	2.0
Perfluorooctanesulfonic acid	2.0 U	2.0
Perfluorooctanoic acid	2.0 U	2.0
Perfluorotetradecanoic acid	2.0 U	2.0
Perfluorotridecanoic acid	2.0 U	2.0
Perfluoroundecanoic acid	2.0 U	2.0
Batch number: 19351014	Sample number(s): 1219261-1219267	
6:2-Fluorotelomersulfonic acid	5.0 U	5.0
8:2-Fluorotelomersulfonic acid	3.0 U	3.0
Perfluorobutanoic acid	5.0 U	5.0
Perfluorodecanesulfonic acid	2.0 U	2.0
Perfluoroheptanesulfonic acid	2.0 U	2.0
Perfluorooctanesulfonamide	2.0 U	2.0
Perfluoropentanoic acid	2.0 U	2.0

LCS/LCSD

Analysis Name	LCS Spike Added	LCS Conc	LCSD Spike Added	LCSD Conc	LCS %REC	LCSD %REC	LCS/LCSD Limits	RPD	RPD Max
	ng/l	ng/l	ng/l	ng/l					
Batch number: 19346028	Sample number(s): 1219261-1219267								
NEtFOSAA	20.48	17.85	20.48	19.44	87	95	70-130	9	30
NMeFOSAA	20.48	18.32	20.48	20.08	89	98	70-130	9	30
Perfluorobutanesulfonic acid	18.12	16.5	18.12	17.8	91	98	70-130	8	30
Perfluorodecanoic acid	20.48	18.9	20.48	20.05	92	98	70-130	6	30
Perfluorododecanoic acid	20.48	18.57	20.48	19.66	91	96	70-130	6	30
Perfluoroheptanoic acid	20.48	19.22	20.48	20.27	94	99	70-130	5	30

*- Outside of specification

(1) The result for one or both determinations was less than five times the LOQ.

(2) The unspiked result was more than four times the spike added.

P##### is indicative of a Background or Unspiked sample that is batch matrix QC and was not performed using a sample from this submission group.

Quality Control Summary

Client Name: C. T. Male Associates
Reported: 12/30/2019 08:44

Group Number: 2078432

LCS/LCSD (continued)

Analysis Name	LCS Spike Added ng/l	LCS Conc ng/l	LCSD Spike Added ng/l	LCSD Conc ng/l	LCS %REC	LCSD %REC	LCS/LCSD Limits	RPD	RPD Max
Perfluorohexanesulfonic acid	18.68	16.9	18.68	18.26	90	98	70-130	8	30
Perfluorohexanoic acid	20.48	18.66	20.48	20.04	91	98	70-130	7	30
Perfluorononanoic acid	20.48	17.81	20.48	19.58	87	96	70-130	10	30
Perfluorooctanesulfonic acid	18.96	17.48	18.96	18.13	92	96	70-130	4	30
Perfluorooctanoic acid	20.48	18.09	20.48	19.14	88	93	70-130	6	30
Perfluorotetradecanoic acid	20.48	18.47	20.48	20.38	90	99	70-130	10	30
Perfluorotridecanoic acid	20.48	19.03	20.48	20.49	93	100	70-130	7	30
Perfluoroundecanoic acid	20.48	19.23	20.48	20.46	94	100	70-130	6	30
Batch number: 19351014	Sample number(s): 1219261-1219267								
6:2-Fluorotelomersulfonic acid	24.28	25.71	24.28	25.17	106	104	56-140	2	30
8:2-Fluorotelomersulfonic acid	24.52	23.14	24.52	24.27	94	99	58-143	5	30
Perfluorobutanoic acid	25.6	24.94	25.6	25.8	97	101	63-160	3	30
Perfluorodecanesulfonic acid	24.64	26.75	24.64	25.24	109	102	62-135	6	30
Perfluoroheptanesulfonic acid	24.36	24.18	24.36	25.14	99	103	67-138	4	30
Perfluorooctanesulfonamide	25.6	24.93	25.6	25.05	97	98	67-126	0	30
Perfluoropentanoic acid	25.6	25.28	25.6	23.8	99	93	73-135	6	30

Surrogate Quality Control

Surrogate recoveries which are outside of the QC window are confirmed unless attributed to dilution or otherwise noted on the Analysis Report.

Analysis Name: 14 PFAS Drinking Water List
Batch number: 19346028

	13C2-PFHxA	13C2-PFDA	D5-NetFOSAA
1219261	92	94	85
1219262	102	103	85
1219263	94	98	86
1219264	97	101	89
1219265	95	100	93
1219266	95	98	87
1219267	95	98	98
Blank	96	104	91
LCS	102	101	93
LCSD	100	104	96
Limits:	70-130	70-130	70-130

Analysis Name: 7 PFAS Compounds
Batch number: 19351014

*- Outside of specification

- (1) The result for one or both determinations was less than five times the LOQ.
- (2) The unspiked result was more than four times the spike added.

P##### is indicative of a Background or Unspiked sample that is batch matrix QC and was not performed using a sample from this submission group.

Quality Control Summary

Client Name: C. T. Male Associates
Reported: 12/30/2019 08:44

Group Number: 2078432

Labeled Isotope Quality Control (continued)

Labeled isotope recoveries which are outside of the QC window are confirmed unless otherwise noted on the analysis report.

Analysis Name: 7 PFAS Compounds
Batch number: 19351014

	13C4-PFBA	13C5-PFPeA	13C3-PFHxS	13C2-6:2-FTS	13C8-PFOS	13C2-8:2-FTS
1219261	97	94	75	110	98	119
1219262	98	110	87	105	104	132
1219263	97	96	70	113	95	112
1219264	90	87	73	113	96	111
1219265	86	85	69	113	88	103
1219266	101	98	77	124	102	123
1219267	90	89	79	119	92	116
Blank	82	81	94	117	100	117
LCS	76	76	90	108	91	115
LCSD	98	104	96	118	104	118
Limits:	43-130	38-150	35-143	29-182	52-121	37-169

	13C8-PFOA
1219261	87
1219262	82
1219263	82
1219264	11
1219265	48
1219266	70
1219267	63
Blank	65
LCS	60
LCSD	91
Limits:	10-134

*- Outside of specification

(1) The result for one or both determinations was less than five times the LOQ.

(2) The unspiked result was more than four times the spike added.

P##### is indicative of a Background or Unspiked sample that is batch matrix QC and was not performed using a sample from this submission group.



Client: C.T Male

Delivery and Receipt Information

Delivery Method:	<u>Fed Ex</u>	Arrival Date:	<u>12/10/2019</u>
Number of Packages:	<u>1</u>	Number of Projects:	<u>1</u>
State/Province of Origin:	<u>NY</u>		

Arrival Condition Summary

Shipping Container Sealed:	Yes	Sample IDs on COC match Containers:	Yes
Custody Seal Present:	Yes	Sample Date/Times match COC:	Yes
Custody Seal Intact:	Yes	Total Trip Blank Qty:	4
Samples Chilled:	Yes	Trip Blank Type:	See Below
Paperwork Enclosed:	Yes	Air Quality Samples Present:	No
Samples Intact:	Yes		
Missing Samples:	No		
Extra Samples:	No		
Discrepancy in Container Qty on COC:	No		

Trip Blank Type(s): 2 Unp, 2 Trizma

Unpacked by Tamara Lugardo

Samples Chilled Details

Thermometer Types: DT = Digital (Temp. Bottle) IR = Infrared (Surface Temp) All Temperatures in °C.

<u>Cooler #</u>	<u>Thermometer ID</u>	<u>Corrected Temp</u>	<u>Therm. Type</u>	<u>Ice Type</u>	<u>Ice Present?</u>	<u>Ice Container</u>	<u>Elevated Temp?</u>
1	192050133	2.5	IR	Wet	Y	Bagged	N

Explanation of Symbols and Abbreviations

The following defines common symbols and abbreviations used in reporting technical data:

BMQL	Below Minimum Quantitation Level	mL	milliliter(s)
C	degrees Celsius	MPN	Most Probable Number
cfu	colony forming units	N.D.	non-detect
CP Units	cobalt-chloroplatinate units	ng	nanogram(s)
F	degrees Fahrenheit	NTU	nephelometric turbidity units
g	gram(s)	pg/L	picogram/liter
IU	International Units	RL	Reporting Limit
kg	kilogram(s)	TNTC	Too Numerous To Count
L	liter(s)	µg	microgram(s)
lb.	pound(s)	µL	microliter(s)
m3	cubic meter(s)	umhos/cm	micromhos/cm
meq	milliequivalents	MCL	Maximum Contamination Limit
mg	milligram(s)		
<	less than		
>	greater than		
ppm	parts per million - One ppm is equivalent to one milligram per kilogram (mg/kg) or one gram per million grams. For aqueous liquids, ppm is usually taken to be equivalent to milligrams per liter (mg/l), because one liter of water has a weight very close to a kilogram. For gases or vapors, one ppm is equivalent to one microliter per liter of gas.		
ppb	parts per billion		
Dry weight basis	Results printed under this heading have been adjusted for moisture content. This increases the analyte weight concentration to approximate the value present in a similar sample without moisture. All other results are reported on an as-received basis.		

Analytical test results meet all requirements of the associated regulatory program (i.e., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis.

Measurement uncertainty values, as applicable, are available upon request.

Tests results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff.

This report shall not be reproduced except in full, without the written approval of the laboratory.

Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" are not performed within 15 minutes.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. THE FOREGOING EXPRESS WARRANTY IS EXCLUSIVE AND IS GIVEN IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED. WE DISCLAIM ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING A WARRANTY OF FITNESS FOR PARTICULAR PURPOSE AND WARRANTY OF MERCHANTABILITY. IN NO EVENT SHALL EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL, LLC BE LIABLE FOR INDIRECT, SPECIAL, CONSEQUENTIAL, OR INCIDENTAL DAMAGES INCLUDING, BUT NOT LIMITED TO, DAMAGES FOR LOSS OF PROFIT OR GOODWILL REGARDLESS OF (A) THE NEGLIGENCE (EITHER SOLE OR CONCURRENT) OF EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL AND (B) WHETHER EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL HAS BEEN INFORMED OF THE POSSIBILITY OF SUCH DAMAGES. We accept no legal responsibility for the purposes for which the client uses the test results. No purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.

Data Qualifiers

Qualifier	Definition
C	Result confirmed by reanalysis
D1	Indicates for dual column analyses that the result is reported from column 1
D2	Indicates for dual column analyses that the result is reported from column 2
E	Concentration exceeds the calibration range
K1	Initial Calibration Blank is above the QC limit and the sample result is ND
K2	Continuing Calibration Blank is above the QC limit and the sample result is ND
K3	Initial Calibration Verification is above the QC limit and the sample result is ND
K4	Continuing Calibration Verification is above the QC limit and the sample result is ND
J (or G, I, X)	Estimated value \geq the Method Detection Limit (MDL or DL) and $<$ the Limit of Quantitation (LOQ or RL)
P	Concentration difference between the primary and confirmation column $>40\%$. The lower result is reported.
P^	Concentration difference between the primary and confirmation column $> 40\%$. The higher result is reported.
U	Analyte was not detected at the value indicated
V	Concentration difference between the primary and confirmation column $>100\%$. The reporting limit is raised due to this disparity and evident interference.
W	The dissolved oxygen uptake for the unseeded blank is greater than 0.20 mg/L.
Z	Laboratory Defined - see analysis report

Additional Organic and Inorganic CLP qualifiers may be used with Form 1 reports as defined by the CLP methods. Qualifiers specific to Dioxin/Furans and PCB Congeners are detailed on the individual Analysis Report.