



ANALYSIS REPORT

Prepared by:

Eurofins Lancaster Laboratories Environmental
2425 New Holland Pike
Lancaster, PA 17601

Prepared for:

C. T. Male Associates
50 Century Hill Drive
Latham NY 12110

Report Date: January 19, 2020 11:35

Project: Hoosick Falls WTP

Account #: 37191
Group Number: 2081402
SDG: HOO41
PO Number: 14.4756
State of Sample Origin: NY

Electronic Copy To	C. T. Male Associates	Attn: Kirk Moline
Electronic Copy To	C. T. Male Associates	Attn: Dan Reilly
Electronic Copy To	C. T. Male Associates	Attn: Jeff Marx
Electronic Copy To	Barr Engineering Company	Attn: Lauren Brady
Electronic Copy To	Environmental Standards	Attn: St. Gobain
Electronic Copy To	Barr Engineering Company	Attn: Data Mgt
Electronic Copy To	Barr Engineering Company	Attn: Terri Olson
Electronic Copy To	C. T. Male Associates	Attn: Nancy Garry

Respectfully Submitted,



(717)-556-7376

To view our laboratory's current scopes of accreditation please go to <https://www.eurofinsus.com/environment-testing/laboratories/eurofins-lancaster-laboratories-environmental/certifications-and-accreditations-eurofins-lancaster-laboratories-environmental/>. Historical copies may be requested through your project manager.



SAMPLE INFORMATION

<u>Client Sample Description</u>	<u>Sample Collection Date/Time</u>	<u>ELLE#</u>
GAC Influent Grab Drinking Water	01/02/2020 09:35	1232645
GAC Midfluent Grab Drinking Water	01/02/2020 09:38	1232646
GAC Effluent Grab Drinking Water	01/02/2020 09:42	1232647
FTB01-200102 Blank Water	01/02/2020 09:45	1232648
LTB01-200102 Blank Water	01/02/2020	1232649

The specific methodologies used in obtaining the enclosed analytical results are indicated on the Laboratory Sample Analysis Record.

Project Name: Hoosick Falls WTP
ELLE Group #: 2081402

General Comments:

See the Laboratory Sample Analysis Record section of the Analysis Report for the method references.

All QC met criteria unless otherwise noted in an Analysis Specific Comment below.

Refer to the QC Summary for specific values and acceptance criteria.

Project specific QC samples are not included in this data set.

Matrix QC may not be reported if site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD was performed, unless otherwise specified in the method.

Surrogate recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in an Analysis Specific Comment below.

The samples were received at the appropriate temperature and in accordance with the chain of custody unless otherwise noted.

Analysis Specific Comments:

No additional comments are necessary.

Sample Description: GAC Influent Grab Drinking Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: PW 1232645
ELLE Group #: 2081402
Matrix: Drinking Water

Project Name: Hoosick Falls WTP

Submission Date/Time: 01/03/2020 10:10
Collection Date/Time: 01/02/2020 09:35
SDG#: HOO41-01

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.8 U	1.8	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.8 U	1.8	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.8 U	1.8	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.8 U	1.8	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.8 U	1.8	1
14070	Perfluoroheptanoic acid¹	375-85-9	15	1.8	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.8 U	1.8	1
14070	Perfluorohexanoic acid¹	307-24-4	16	1.8	1
14070	Perfluorononanoic acid ¹	375-95-1	1.8 U	1.8	1
14070	Perfluorooctanesulfonic acid¹	1763-23-1	3.3	1.8	1
14070	Perfluorooctanoic acid¹	335-67-1	570	18	10
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.8 U	1.8	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.8 U	1.8	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.8 U	1.8	1

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified					ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.5 U	4.5	1		
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1		
14473	Perfluorobutanoic acid¹	375-22-4	7.0	6.3	1		
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1		
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1		
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1		
14473	Perfluoropentanoic acid¹	2706-90-3	4.4	1.8	1		

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	20003013	01/08/2020 03:00	Marissa C Drexinger	1
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	20003013	01/10/2020 21:26	Marissa C Drexinger	10
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	20003014	01/09/2020 01:44	Christine E Dolman	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	20003013	01/05/2020 18:00	Anthony C Polaski	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	20003014	01/05/2020 18:00	Anthony C Polaski	1

Sample Description: GAC Midfluent Grab Drinking Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: PW 1232646
ELLE Group #: 2081402
Matrix: Drinking Water

Project Name: Hoosick Falls WTP

Submission Date/Time: 01/03/2020 10:10
Collection Date/Time: 01/02/2020 09:38
SDG#: HOO41-02

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.8 U	1.8	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.8 U	1.8	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.8 U	1.8	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.8 U	1.8	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.8 U	1.8	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.8 U	1.8	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.8 U	1.8	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.8 U	1.8	1
14070	Perfluorononanoic acid ¹	375-95-1	1.8 U	1.8	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.8 U	1.8	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.8 U	1.8	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.8 U	1.8	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.8 U	1.8	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.8 U	1.8	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.4 U	4.4	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.6 U	2.6	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.1 U	6.1	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.7 U	1.7	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.7 U	1.7	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.7 U	1.7	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.7 U	1.7	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	20003013	01/08/2020 03:11	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	20003014	01/09/2020 01:53	Christine E Dolman	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	20003013	01/05/2020 18:00	Anthony C Polaski	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	20003014	01/05/2020 18:00	Anthony C Polaski	1

Sample Description: GAC Effluent Grab Drinking Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: PW 1232647
ELLE Group #: 2081402
Matrix: Drinking Water

Project Name: Hoosick Falls WTP

Submittal Date/Time: 01/03/2020 10:10
Collection Date/Time: 01/02/2020 09:42
SDG#: HOO41-03

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1					
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.8 U	1.8	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.8 U	1.8	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.8 U	1.8	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.8 U	1.8	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.8 U	1.8	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.8 U	1.8	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.8 U	1.8	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.8 U	1.8	1
14070	Perfluorononanoic acid ¹	375-95-1	1.8 U	1.8	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.8 U	1.8	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.8 U	1.8	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.8 U	1.8	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.8 U	1.8	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.8 U	1.8	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified					
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.4 U	4.4	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.6 U	2.6	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.2 U	6.2	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	20003013	01/08/2020 03:23	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	20003014	01/09/2020 02:02	Christine E Dolman	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	20003013	01/05/2020 18:00	Anthony C Polaski	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	20003014	01/05/2020 18:00	Anthony C Polaski	1

Sample Description: FTB01-200102 Blank Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: WW 1232648
ELLE Group #: 2081402
Matrix: Blank Water

Project Name: Hoosick Falls WTP

Submittal Date/Time: 01/03/2020 10:10
Collection Date/Time: 01/02/2020 09:45
SDG#: HOO41-04FB

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1			ng/l	ng/l	
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.8 U	1.8	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.8 U	1.8	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.8 U	1.8	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.8 U	1.8	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.8 U	1.8	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.8 U	1.8	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.8 U	1.8	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.8 U	1.8	1
14070	Perfluorononanoic acid ¹	375-95-1	1.8 U	1.8	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.8 U	1.8	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.8 U	1.8	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.8 U	1.8	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.8 U	1.8	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.8 U	1.8	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified			ng/l	ng/l	
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.5 U	4.5	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.3 U	6.3	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	20003013	01/08/2020 03:34	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	20003014	01/09/2020 02:11	Christine E Dolman	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	20003013	01/05/2020 18:00	Anthony C Polaski	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	20003014	01/05/2020 18:00	Anthony C Polaski	1

Sample Description: LTB01-200102 Blank Water
Hoosick Falls Water Treatment Plant

C. T. Male Associates
ELLE Sample #: WW 1232649
ELLE Group #: 2081402
Matrix: Blank Water

Project Name: Hoosick Falls WTP

Submittal Date/Time: 01/03/2020 10:10
Collection Date/Time: 01/02/2020
SDG#: HOO41-05TB

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1					
14070	NEtFOSAA ¹ NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.	2991-50-6	1.8 U	1.8	1
14070	NMeFOSAA ¹ NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.	2355-31-9	1.8 U	1.8	1
14070	Perfluorobutanesulfonic acid ¹	375-73-5	1.8 U	1.8	1
14070	Perfluorodecanoic acid ¹	335-76-2	1.8 U	1.8	1
14070	Perfluorododecanoic acid ¹	307-55-1	1.8 U	1.8	1
14070	Perfluoroheptanoic acid ¹	375-85-9	1.8 U	1.8	1
14070	Perfluorohexanesulfonic acid ¹	355-46-4	1.8 U	1.8	1
14070	Perfluorohexanoic acid ¹	307-24-4	1.8 U	1.8	1
14070	Perfluorononanoic acid ¹	375-95-1	1.8 U	1.8	1
14070	Perfluorooctanesulfonic acid ¹	1763-23-1	1.8 U	1.8	1
14070	Perfluorooctanoic acid ¹	335-67-1	1.8 U	1.8	1
14070	Perfluorotetradecanoic acid ¹	376-06-7	1.8 U	1.8	1
14070	Perfluorotridecanoic acid ¹	72629-94-8	1.8 U	1.8	1
14070	Perfluoroundecanoic acid ¹	2058-94-8	1.8 U	1.8	1

CAT No.	Analysis Name	CAS Number	Result	Limit of Quantitation	Dilution Factor
LC/MS/MS Miscellaneous EPA 537 Version 1.1 Modified					
14473	6:2-Fluorotelomersulfonic acid ¹	27619-97-2	4.5 U	4.5	1
14473	8:2-Fluorotelomersulfonic acid ¹	39108-34-4	2.7 U	2.7	1
14473	Perfluorobutanoic acid ¹	375-22-4	6.2 U	6.2	1
14473	Perfluorodecanesulfonic acid ¹	335-77-3	1.8 U	1.8	1
14473	Perfluoroheptanesulfonic acid ¹	375-92-8	1.8 U	1.8	1
14473	Perfluorooctanesulfonamide ¹	754-91-6	1.8 U	1.8	1
14473	Perfluoropentanoic acid ¹	2706-90-3	1.8 U	1.8	1

Sample Comments

¹ = This analyte was not on the laboratory's NYSDOH Scope of Accreditation at the time of analysis.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	14 PFAS Drinking Water List	EPA 537 Version 1.1	1	20003013	01/08/2020 03:46	Marissa C Drexinger	1
14473	7 PFAS Compounds	EPA 537 Version 1.1 Modified	1	20003014	01/09/2020 02:29	Christine E Dolman	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	20003013	01/05/2020 18:00	Anthony C Polaski	1
14091	PFAS Water Prep	EPA 537 Version 1.1 Modified	1	20003014	01/05/2020 18:00	Anthony C Polaski	1

Quality Control Summary

Client Name: C. T. Male Associates
Reported: 01/19/2020 11:35

Group Number: 2081402

Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD was performed, unless otherwise specified in the method.

All Inorganic Initial Calibration and Continuing Calibration Blanks met acceptable method criteria unless otherwise noted on the Analysis Report.

Method Blank

Analysis Name	Result	LOQ
	ng/l	ng/l
Batch number: 20003013	Sample number(s): 1232645-1232649	
NEtFOSAA	2.0 U	2.0
NMeFOSAA	2.0 U	2.0
Perfluorobutanesulfonic acid	2.0 U	2.0
Perfluorodecanoic acid	2.0 U	2.0
Perfluorododecanoic acid	2.0 U	2.0
Perfluoroheptanoic acid	2.0 U	2.0
Perfluorohexanesulfonic acid	2.0 U	2.0
Perfluorohexanoic acid	2.0 U	2.0
Perfluorononanoic acid	2.0 U	2.0
Perfluorooctanesulfonic acid	2.0 U	2.0
Perfluorooctanoic acid	2.0 U	2.0
Perfluorotetradecanoic acid	2.0 U	2.0
Perfluorotridecanoic acid	2.0 U	2.0
Perfluoroundecanoic acid	2.0 U	2.0
Batch number: 20003014	Sample number(s): 1232645-1232649	
6:2-Fluorotelomersulfonic acid	5.0 U	5.0
8:2-Fluorotelomersulfonic acid	3.0 U	3.0
Perfluorobutanoic acid	5.0 U	5.0
Perfluorodecanesulfonic acid	2.0 U	2.0
Perfluoroheptanesulfonic acid	2.0 U	2.0
Perfluorooctanesulfonamide	2.0 U	2.0
Perfluoropentanoic acid	2.0 U	2.0

LCS/LCSD

Analysis Name	LCS Spike Added	LCS Conc	LCSD Spike Added	LCSD Conc	LCS %REC	LCSD %REC	LCS/LCSD Limits	RPD	RPD Max
	ng/l	ng/l	ng/l	ng/l					
Batch number: 20003013	Sample number(s): 1232645-1232649								
NEtFOSAA	20.48	21.13	20.48	20.05	103	98	70-130	5	30
NMeFOSAA	20.48	20.42	20.48	20.64	100	101	70-130	1	30
Perfluorobutanesulfonic acid	18.12	17.38	18.12	17.86	96	99	70-130	3	30
Perfluorodecanoic acid	20.48	20.56	20.48	22.27	100	109	70-130	8	30
Perfluorododecanoic acid	20.48	21.15	20.48	21.84	103	107	70-130	3	30
Perfluoroheptanoic acid	20.48	20.72	20.48	20.93	101	102	70-130	1	30

*- Outside of specification

(1) The result for one or both determinations was less than five times the LOQ.

(2) The unspiked result was more than four times the spike added.

P##### is indicative of a Background or Unspiked sample that is batch matrix QC and was not performed using a sample from this submission group.

Quality Control Summary

Client Name: C. T. Male Associates
Reported: 01/19/2020 11:35

Group Number: 2081402

LCS/LCSD (continued)

Analysis Name	LCS Spike Added ng/l	LCS Conc ng/l	LCSD Spike Added ng/l	LCSD Conc ng/l	LCS %REC	LCSD %REC	LCS/LCSD Limits	RPD	RPD Max
Perfluorohexanesulfonic acid	18.68	18.99	18.68	19.01	102	102	70-130	0	30
Perfluorohexanoic acid	20.48	19.99	20.48	21.17	98	103	70-130	6	30
Perfluorononanoic acid	20.48	19.71	20.48	21.33	96	104	70-130	8	30
Perfluorooctanesulfonic acid	18.96	18.55	18.96	18.78	98	99	70-130	1	30
Perfluorooctanoic acid	20.48	20.29	20.48	21.52	99	105	70-130	6	30
Perfluorotetradecanoic acid	20.48	19.77	20.48	20.46	97	100	70-130	3	30
Perfluorotridecanoic acid	20.48	19.99	20.48	21.53	98	105	70-130	7	30
Perfluoroundecanoic acid	20.48	21.66	20.48	23.07	106	113	70-130	6	30
Batch number: 20003014	Sample number(s): 1232645-1232649								
6:2-Fluorotelomersulfonic acid	24.28	21.67	24.28	22.27	89	92	56-140	3	30
8:2-Fluorotelomersulfonic acid	24.52	21.66	24.52	20.88	88	85	58-143	4	30
Perfluorobutanoic acid	25.6	23.44	25.6	24.52	92	96	63-160	5	30
Perfluorodecanesulfonic acid	24.64	22.61	24.64	20.27	92	82	62-135	11	30
Perfluoroheptanesulfonic acid	24.36	22.46	24.36	21.91	92	90	67-138	2	30
Perfluorooctanesulfonamide	25.6	20.63	25.6	22.26	81	87	67-126	8	30
Perfluoropentanoic acid	25.6	22.23	25.6	24.25	87	95	73-135	9	30

Surrogate Quality Control

Surrogate recoveries which are outside of the QC window are confirmed unless attributed to dilution or otherwise noted on the Analysis Report.

Analysis Name: 14 PFAS Drinking Water List
Batch number: 20003013

	13C2-PFHxA	13C2-PFDA	D5-NetFOSAA
1232645	110	105	95
1232646	98	93	90
1232647	97	92	89
1232648	89	86	83
1232649	101	102	96
Blank	98	103	98
LCS	93	93	93
LCSD	100	103	96
Limits:	70-130	70-130	70-130

Analysis Name: 7 PFAS Compounds
Batch number: 20003014

	13C4-PFBA	13C5-PFPeA	13C3-PFHxS	13C2-6:2-FTS	13C8-PFOS	13C2-8:2-FTS
1232645	107	124	137	117	110	131
1232646	101	103	114	120	104	117

*- Outside of specification

- (1) The result for one or both determinations was less than five times the LOQ.
- (2) The unspiked result was more than four times the spike added.

P##### is indicative of a Background or Unspiked sample that is batch matrix QC and was not performed using a sample from this submission group.

Quality Control Summary

Client Name: C. T. Male Associates
Reported: 01/19/2020 11:35

Group Number: 2081402

Labeled Isotope Quality Control (continued)

Labeled isotope recoveries which are outside of the QC window are confirmed unless otherwise noted on the analysis report.

Analysis Name: 7 PFAS Compounds
Batch number: 20003014

	13C4-PFBA	13C5-PFPeA	13C3-PFHxS	13C2-6:2-FTS	13C8-PFOS	13C2-8:2-FTS
1232647	107	102	114	116	100	113
1232648	106	104	125	118	107	121
1232649	110	109	112	115	107	122
Blank	107	106	111	120	101	132
LCS	103	104	104	108	95	104
LCSD	104	101	115	114	106	106
Limits:	43-130	38-150	35-143	29-182	52-121	37-169

	13C8-PFOA
1232645	48
1232646	100
1232647	70
1232648	105
1232649	115
Blank	112
LCS	96
LCSD	96
Limits:	10-134

*- Outside of specification

(1) The result for one or both determinations was less than five times the LOQ.

(2) The unspiked result was more than four times the spike added.

P##### is indicative of a Background or Unspiked sample that is batch matrix QC and was not performed using a sample from this submission group.

Environmental Analysis Request/Chain of Custody



Lancaster Laboratories Environmental

Acct. # 37191 Group # 2001402 Sample # 1232645-99

Client: <u>C-T, Mall Associates</u>				Matrix				Analyses Requested								For Lab Use Only										
Project Name/#: <u>HOOSICK FALLS WTP</u>								Site ID #: _____				Preservation and Filtration Codes								SF #: _____						
Project Manager: <u>Kirk Moline</u>				P.O. #: <u>14 4756</u>				<input type="checkbox"/> Soil	<input type="checkbox"/> Tissue	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	SCR #: _____						
Sampler: <u>Kendall Costas</u>				PWSID #: _____				<input type="checkbox"/> Potable	<input type="checkbox"/> Ground	<input type="checkbox"/> Surface																
Phone #: _____				Quote #: _____				<input type="checkbox"/> NPDES	<input type="checkbox"/>	<input type="checkbox"/>																
State where samples were collected: <u>NY</u>				For Compliance: Yes <input type="checkbox"/> No <input type="checkbox"/>				Other: <u>reagent water</u>				Total # of Containers								Preservation Codes H = HCl T = Thiosulfate N = HNO ₃ B = NaOH S = H ₂ SO ₄ P = H ₃ PO ₄ F = Field Filtered O = Other						
Sample Identification				Collection		<input type="checkbox"/> Sediment	<input type="checkbox"/>															<input type="checkbox"/>	7 PPAAS (EPA method 837)			
	Date	Time	Grab	Composite	<input type="checkbox"/> Water																					
<u>GAC Influent</u>	<u>1/2/20</u>	<u>935</u>	<input checked="" type="checkbox"/>				<input checked="" type="checkbox"/>														<u>2 Tr, 2 Env, 2 NO preservative</u>					
<u>GAC Midfluent</u>	<u>↓</u>	<u>938</u>	<input checked="" type="checkbox"/>				<input checked="" type="checkbox"/>																			
<u>GAC Effluent</u>	<u>↓</u>	<u>942</u>	<input checked="" type="checkbox"/>				<input checked="" type="checkbox"/>																			
<u>FTBD1 -200102</u>	<u>1/2/20</u>	<u>945</u>	<input checked="" type="checkbox"/>						<input checked="" type="checkbox"/>																	
<u>LTB01 -200102</u>	<u>↓</u>	<u>-</u>	<input checked="" type="checkbox"/>						<input checked="" type="checkbox"/>																	
Turnaround Time Requested (TAT) (please check): Standard <input checked="" type="checkbox"/> Rush <input type="checkbox"/>				Relinquished by: <u>Kendall Costas</u>				Date	Time	Received by:		Date	Time													
(Rush TAT is subject to laboratory approval and surcharges.)								<u>1/2/20</u>	<u>1050</u>																	
Date results are needed:				Relinquished by:				Date	Time	Received by:		Date	Time													
Rush results requested by (please check): E-Mail <input type="checkbox"/> Phone <input type="checkbox"/>																										
E-mail Address:				Relinquished by:				Date	Time	Received by:		Date	Time													
Phone:																										
Data Package Options (please check if required)				Relinquished by:				Date	Time	Received by:		Date	Time													
Type I (Validation/non-CLP) <input type="checkbox"/>	MA MCP <input type="checkbox"/>																									
Type III (Reduced non-CLP) <input type="checkbox"/>	CT RCP <input type="checkbox"/>																									
Type VI (Raw Data Only) <input type="checkbox"/>	TX TRRP-13 <input type="checkbox"/>																									
NJ DKQP <input type="checkbox"/>	NYSDEC Category <input type="checkbox"/> A or <input checked="" type="checkbox"/> B																									
EDD Required? Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> If yes, format: <u>EQMIS</u>				Relinquished by Commercial Carrier:																						
				UPS <input type="checkbox"/> FedEx <input checked="" type="checkbox"/> Other <input type="checkbox"/>								<u>1.2</u>														
												Temperature upon receipt														



Client: C.T. MALE ASSOCIATES

Delivery and Receipt Information

Delivery Method: Fed Ex Arrival Date: 01/03/2020
 Number of Packages: 1 Number of Projects: 1
 State/Province of Origin: New York

Arrival Condition Summary

Shipping Container Sealed:	Yes	Sample IDs on COC match Containers:	Yes
Custody Seal Present:	Yes	Sample Date/Times match COC:	Yes
Custody Seal Intact:	Yes	Total Trip Blank Qty:	4
Samples Chilled:	Yes	Trip Blank Type:	See Below
Paperwork Enclosed:	Yes	Air Quality Samples Present:	No
Samples Intact:	Yes		
Missing Samples:	No		
Extra Samples:	No		
Discrepancy in Container Qty on COC:	No		

Trip Blank Type(s): 2 TRIZMA / 2 UNP

Unpacked by *Julissa Rivera-Santa*

Samples Chilled Details

Thermometer Types: *DT = Digital (Temp. Bottle)* *IR = Infrared (Surface Temp)* All Temperatures in °C.

Cooler #	Thermometer ID	Corrected Temp	Therm. Type	Ice Type	Ice Present?	Ice Container	Elevated Temp?
1	192050133	1.2	IR	Wet	Y	Bagged	N

Explanation of Symbols and Abbreviations

The following defines common symbols and abbreviations used in reporting technical data:

BMQL	Below Minimum Quantitation Level	mL	milliliter(s)
C	degrees Celsius	MPN	Most Probable Number
cfu	colony forming units	N.D.	non-detect
CP Units	cobalt-chloroplatinate units	ng	nanogram(s)
F	degrees Fahrenheit	NTU	nephelometric turbidity units
g	gram(s)	pg/L	picogram/liter
IU	International Units	RL	Reporting Limit
kg	kilogram(s)	TNTC	Too Numerous To Count
L	liter(s)	µg	microgram(s)
lb.	pound(s)	µL	microliter(s)
m3	cubic meter(s)	umhos/cm	micromhos/cm
meq	milliequivalents	MCL	Maximum Contamination Limit
mg	milligram(s)		
<	less than		
>	greater than		
ppm	parts per million - One ppm is equivalent to one milligram per kilogram (mg/kg) or one gram per million grams. For aqueous liquids, ppm is usually taken to be equivalent to milligrams per liter (mg/l), because one liter of water has a weight very close to a kilogram. For gases or vapors, one ppm is equivalent to one microliter per liter of gas.		
ppb	parts per billion		
Dry weight basis	Results printed under this heading have been adjusted for moisture content. This increases the analyte weight concentration to approximate the value present in a similar sample without moisture. All other results are reported on an as-received basis.		

Analytical test results meet all requirements of the associated regulatory program (i.e., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis.

Measurement uncertainty values, as applicable, are available upon request.

Tests results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff.

This report shall not be reproduced except in full, without the written approval of the laboratory.

Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" are not performed within 15 minutes.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. THE FOREGOING EXPRESS WARRANTY IS EXCLUSIVE AND IS GIVEN IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED. WE DISCLAIM ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING A WARRANTY OF FITNESS FOR PARTICULAR PURPOSE AND WARRANTY OF MERCHANTABILITY. IN NO EVENT SHALL EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL, LLC BE LIABLE FOR INDIRECT, SPECIAL, CONSEQUENTIAL, OR INCIDENTAL DAMAGES INCLUDING, BUT NOT LIMITED TO, DAMAGES FOR LOSS OF PROFIT OR GOODWILL REGARDLESS OF (A) THE NEGLIGENCE (EITHER SOLE OR CONCURRENT) OF EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL AND (B) WHETHER EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL HAS BEEN INFORMED OF THE POSSIBILITY OF SUCH DAMAGES. We accept no legal responsibility for the purposes for which the client uses the test results. No purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.

Data Qualifiers

Qualifier	Definition
C	Result confirmed by reanalysis
D1	Indicates for dual column analyses that the result is reported from column 1
D2	Indicates for dual column analyses that the result is reported from column 2
E	Concentration exceeds the calibration range
K1	Initial Calibration Blank is above the QC limit and the sample result is ND
K2	Continuing Calibration Blank is above the QC limit and the sample result is ND
K3	Initial Calibration Verification is above the QC limit and the sample result is ND
K4	Continuing Calibration Verification is above the QC limit and the sample result is ND
J (or G, I, X)	Estimated value \geq the Method Detection Limit (MDL or DL) and $<$ the Limit of Quantitation (LOQ or RL)
P	Concentration difference between the primary and confirmation column $>40\%$. The lower result is reported.
P^	Concentration difference between the primary and confirmation column $> 40\%$. The higher result is reported.
U	Analyte was not detected at the value indicated
V	Concentration difference between the primary and confirmation column $>100\%$. The reporting limit is raised due to this disparity and evident interference.
W	The dissolved oxygen uptake for the unseeded blank is greater than 0.20 mg/L.
Z	Laboratory Defined - see analysis report

Additional Organic and Inorganic CLP qualifiers may be used with Form 1 reports as defined by the CLP methods. Qualifiers specific to Dioxin/Furans and PCB Congeners are detailed on the individual Analysis Report.