


ANALYTICAL REPORT

Eurofins Lancaster Laboratories Env, LLC
2425 New Holland Pike
Lancaster, PA 17601
Tel: (717)656-2300

Laboratory Job ID: 410-71774-1
Laboratory Sample Delivery Group: HOO
Client Project/Site: Hoosick Falls WTP

For:
CT Male Associates DPC
50 Century Hill Dr
Latham, New York 12110

Attn: Mr. Kirk Moline



Authorized for release by:
2/17/2022 12:43:09 PM

Paul Hobart, Project Manager
(617)312-8660
Paul.Hobart@Eurofinset.com

LINKS

Review your project
results through
TotalAccess

Have a Question?



Visit us at:

www.eurofinsus.com/Env

The test results in this report meet all 2003 NELAC, 2009 TNI, and 2016 TNI requirements for accredited parameters, exceptions are noted in this report. This report may not be reproduced except in full, and with written approval from the laboratory. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.



Analytical test results meet all requirements of the associated regulatory program (e.g., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis. Data qualifiers are applied to note exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- QC results that exceed the upper limits and are associated with non-detect samples are qualified but further narration is not required since the bias is high and does not change a non-detect result. Further narration is also not required with QC blank detection when the associated sample concentration is non-detect or more than ten times the level in the blank.
 - Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD is performed, unless otherwise specified in the method.
 - Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.
- Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Measurement uncertainty values, as applicable, are available upon request.

Test results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" and tested in the laboratory are not performed within 15 minutes of collection.

This report shall not be reproduced except in full, without the written approval of the laboratory.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. The foregoing express warranty is exclusive and is given in lieu of all other warranties, expressed or implied, except as otherwise agreed. We disclaim any other warranties, expressed or implied, including a warranty of fitness for particular purpose and warranty of merchantability. In no event shall Eurofins Lancaster Laboratories Environmental, LLC be liable for indirect, special, consequential, or incidental damages including, but not limited to, damages for loss of profit or goodwill regardless of (A) the negligence (either sole or concurrent) of Eurofins Lancaster Laboratories Environmental and (B) whether Eurofins Lancaster Laboratories Environmental has been informed of the possibility of such damages. We accept no legal responsibility for the purposes for which the client uses the test results. Except as otherwise agreed, no purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.

A handwritten signature in black ink, appearing to read "Paul Hobart".

Paul Hobart
Project Manager
2/17/2022 12:43:09 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Surrogate Summary	13
Isotope Dilution Summary	14
QC Sample Results	15
QC Association Summary	18
Lab Chronicle	19
Certification Summary	21
Method Summary	22
Sample Summary	23
Chain of Custody	24
Receipt Checklists	25

Definitions/Glossary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Qualifiers

LCMS

Qualifier	Qualifier Description
U	Indicates the analyte was analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
1C	Result is from the primary column on a dual-column method.
2C	Result is from the confirmation column on a dual-column method.
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Job ID: 410-71774-1

Laboratory: Eurofins Lancaster Laboratories Env, LLC

Narrative

Job Narrative
410-71774-1

Receipt

The samples were received on 2/4/2022 10:01 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 0.5°C

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Detection Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: GAC Influent

Lab Sample ID: 410-71774-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorooctanesulfonamide	3.7		1.8	ng/L	1		537 (Mod)	Total/NA
Perfluoropentanoic acid	3.0		1.8	ng/L	1		537 (Mod)	Total/NA
Perfluorohexanoic acid	12		1.8	ng/L	1		537 DW	Total/NA
Perfluoroheptanoic acid	13		1.8	ng/L	1		537 DW	Total/NA
Perfluorooctanesulfonic acid	3.3		1.8	ng/L	1		537 DW	Total/NA
Perfluorooctanoic acid - DL	460		18	ng/L	10		537 DW	Total/NA

Client Sample ID: GAC Midfluent

Lab Sample ID: 410-71774-2

No Detections.

Client Sample ID: GAC Effluent

Lab Sample ID: 410-71774-3

No Detections.

Client Sample ID: PV-2_25

Lab Sample ID: 410-71774-4

No Detections.

Client Sample ID: FTB01-220203

Lab Sample ID: 410-71774-5

No Detections.

Client Sample ID: LTB01-220203

Lab Sample ID: 410-71774-6

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Env, LLC

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: GAC Influent

Lab Sample ID: 410-71774-1

Date Collected: 02/03/22 14:40

Matrix: Water

Date Received: 02/04/22 10:01

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	4.4	U	4.4	ng/L		02/15/22 07:57	02/17/22 00:50	1
8:2 Fluorotelomer sulfonic acid	2.6	U	2.6	ng/L		02/15/22 07:57	02/17/22 00:50	1
Perfluorobutanoic acid	4.4	U	4.4	ng/L		02/15/22 07:57	02/17/22 00:50	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		02/15/22 07:57	02/17/22 00:50	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		02/15/22 07:57	02/17/22 00:50	1
Perfluorooctanesulfonamide	3.7		1.8	ng/L		02/15/22 07:57	02/17/22 00:50	1
Perfluoropentanoic acid	3.0		1.8	ng/L		02/15/22 07:57	02/17/22 00:50	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	89		17 - 200			02/15/22 07:57	02/17/22 00:50	1
M2-8:2 FTS	105		33 - 200			02/15/22 07:57	02/17/22 00:50	1
13C4 PFBA	108		42 - 165			02/15/22 07:57	02/17/22 00:50	1
13C5 PFPeA	111		38 - 187			02/15/22 07:57	02/17/22 00:50	1
13C8 PFOS	104		51 - 159			02/15/22 07:57	02/17/22 00:50	1
13C8 FOSA	64		10 - 168			02/15/22 07:57	02/17/22 00:50	1
13C3 PFHxS	113		28 - 188			02/15/22 07:57	02/17/22 00:50	1

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	12		1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluoroheptanoic acid	13		1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluorooctanesulfonic acid	3.3		1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
NEtFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
NMeFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:53	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	79		70 - 130			02/10/22 17:14	02/14/22 15:53	1
13C2 PFDA	104		70 - 130			02/10/22 17:14	02/14/22 15:53	1
13C2 PFHxA	112		70 - 130			02/10/22 17:14	02/14/22 15:53	1

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS) - DL

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanoic acid	460		18	ng/L		02/10/22 17:14	02/15/22 09:53	10
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	86		70 - 130			02/10/22 17:14	02/15/22 09:53	10
13C2 PFDA	90		70 - 130			02/10/22 17:14	02/15/22 09:53	10
13C2 PFHxA	108		70 - 130			02/10/22 17:14	02/15/22 09:53	10

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: GAC Midfluent

Lab Sample ID: 410-71774-2

Date Collected: 02/03/22 14:45

Matrix: Water

Date Received: 02/04/22 10:01

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	4.3	U	4.3	ng/L		02/15/22 07:57	02/17/22 01:01	1
8:2 Fluorotelomer sulfonic acid	2.6	U	2.6	ng/L		02/15/22 07:57	02/17/22 01:01	1
Perfluorobutanoic acid	4.3	U	4.3	ng/L		02/15/22 07:57	02/17/22 01:01	1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:01	1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:01	1
Perfluorooctanesulfonamide	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:01	1
Perfluoropentanoic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:01	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	112		17 - 200	02/15/22 07:57	02/17/22 01:01	1
M2-8:2 FTS	111		33 - 200	02/15/22 07:57	02/17/22 01:01	1
13C4 PFBA	112		42 - 165	02/15/22 07:57	02/17/22 01:01	1
13C5 PFPeA	119		38 - 187	02/15/22 07:57	02/17/22 01:01	1
13C8 PFOS	115		51 - 159	02/15/22 07:57	02/17/22 01:01	1
13C8 FOSA	85		10 - 168	02/15/22 07:57	02/17/22 01:01	1
13C3 PFHxS	110		28 - 188	02/15/22 07:57	02/17/22 01:01	1

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
NEtFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
NMeFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 14:55	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	84		70 - 130	02/10/22 17:14	02/14/22 14:55	1
13C2 PFDA	80		70 - 130	02/10/22 17:14	02/14/22 14:55	1
13C2 PFHxA	94		70 - 130	02/10/22 17:14	02/14/22 14:55	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: GAC Effluent

Lab Sample ID: 410-71774-3

Date Collected: 02/03/22 14:48

Matrix: Water

Date Received: 02/04/22 10:01

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	4.2	U	4.2	ng/L		02/15/22 07:57	02/17/22 01:12	1
8:2 Fluorotelomer sulfonic acid	2.5	U	2.5	ng/L		02/15/22 07:57	02/17/22 01:12	1
Perfluorobutanoic acid	4.2	U	4.2	ng/L		02/15/22 07:57	02/17/22 01:12	1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:12	1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:12	1
Perfluorooctanesulfonamide	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:12	1
Perfluoropentanoic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:12	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	114		17 - 200	02/15/22 07:57	02/17/22 01:12	1
M2-8:2 FTS	112		33 - 200	02/15/22 07:57	02/17/22 01:12	1
13C4 PFBA	114		42 - 165	02/15/22 07:57	02/17/22 01:12	1
13C5 PFPeA	108		38 - 187	02/15/22 07:57	02/17/22 01:12	1
13C8 PFOS	113		51 - 159	02/15/22 07:57	02/17/22 01:12	1
13C8 FOSA	82		10 - 168	02/15/22 07:57	02/17/22 01:12	1
13C3 PFHxS	100		28 - 188	02/15/22 07:57	02/17/22 01:12	1

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
NEtFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
NMeFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:06	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	85		70 - 130	02/10/22 17:14	02/14/22 15:06	1
13C2 PFDA	84		70 - 130	02/10/22 17:14	02/14/22 15:06	1
13C2 PFHxA	95		70 - 130	02/10/22 17:14	02/14/22 15:06	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: PV-2_25

Lab Sample ID: 410-71774-4

Date Collected: 02/03/22 14:50

Matrix: Water

Date Received: 02/04/22 10:01

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	4.3	U	4.3	ng/L		02/15/22 07:57	02/17/22 01:23	1
8:2 Fluorotelomer sulfonic acid	2.6	U	2.6	ng/L		02/15/22 07:57	02/17/22 01:23	1
Perfluorobutanoic acid	4.3	U	4.3	ng/L		02/15/22 07:57	02/17/22 01:23	1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:23	1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:23	1
Perfluorooctanesulfonamide	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:23	1
Perfluoropentanoic acid	1.7	U	1.7	ng/L		02/15/22 07:57	02/17/22 01:23	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	110		17 - 200	02/15/22 07:57	02/17/22 01:23	1
M2-8:2 FTS	112		33 - 200	02/15/22 07:57	02/17/22 01:23	1
13C4 PFBA	115		42 - 165	02/15/22 07:57	02/17/22 01:23	1
13C5 PFPeA	110		38 - 187	02/15/22 07:57	02/17/22 01:23	1
13C8 PFOS	110		51 - 159	02/15/22 07:57	02/17/22 01:23	1
13C8 FOSA	76		10 - 168	02/15/22 07:57	02/17/22 01:23	1
13C3 PFHxS	98		28 - 188	02/15/22 07:57	02/17/22 01:23	1

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
NEtFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
NMeFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:18	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	85		70 - 130	02/10/22 17:14	02/14/22 15:18	1
13C2 PFDA	81		70 - 130	02/10/22 17:14	02/14/22 15:18	1
13C2 PFHxA	95		70 - 130	02/10/22 17:14	02/14/22 15:18	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: FTB01-220203

Lab Sample ID: 410-71774-5

Date Collected: 02/03/22 14:55

Matrix: Water

Date Received: 02/04/22 10:01

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	4.6	U	4.6	ng/L		02/15/22 07:57	02/17/22 01:34	1
8:2 Fluorotelomer sulfonic acid	2.7	U	2.7	ng/L		02/15/22 07:57	02/17/22 01:34	1
Perfluorobutanoic acid	4.6	U	4.6	ng/L		02/15/22 07:57	02/17/22 01:34	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		02/15/22 07:57	02/17/22 01:34	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		02/15/22 07:57	02/17/22 01:34	1
Perfluorooctanesulfonamide	1.8	U	1.8	ng/L		02/15/22 07:57	02/17/22 01:34	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		02/15/22 07:57	02/17/22 01:34	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	104		17 - 200	02/15/22 07:57	02/17/22 01:34	1
M2-8:2 FTS	116		33 - 200	02/15/22 07:57	02/17/22 01:34	1
13C4 PFBA	116		42 - 165	02/15/22 07:57	02/17/22 01:34	1
13C5 PFPeA	111		38 - 187	02/15/22 07:57	02/17/22 01:34	1
13C8 PFOS	116		51 - 159	02/15/22 07:57	02/17/22 01:34	1
13C8 FOSA	83		10 - 168	02/15/22 07:57	02/17/22 01:34	1
13C3 PFHxS	99		28 - 188	02/15/22 07:57	02/17/22 01:34	1

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
NEtFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
NMeFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:29	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	83		70 - 130	02/10/22 17:14	02/14/22 15:29	1
13C2 PFDA	82		70 - 130	02/10/22 17:14	02/14/22 15:29	1
13C2 PFHxA	91		70 - 130	02/10/22 17:14	02/14/22 15:29	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: LTB01-220203

Lab Sample ID: 410-71774-6

Date Collected: 02/03/22 00:00

Matrix: Water

Date Received: 02/04/22 10:01

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	4.7	U	4.7	ng/L		02/15/22 07:57	02/17/22 01:56	1
8:2 Fluorotelomer sulfonic acid	2.8	U	2.8	ng/L		02/15/22 07:57	02/17/22 01:56	1
Perfluorobutanoic acid	4.7	U	4.7	ng/L		02/15/22 07:57	02/17/22 01:56	1
Perfluorodecanesulfonic acid	1.9	U	1.9	ng/L		02/15/22 07:57	02/17/22 01:56	1
Perfluoroheptanesulfonic acid	1.9	U	1.9	ng/L		02/15/22 07:57	02/17/22 01:56	1
Perfluorooctanesulfonamide	1.9	U	1.9	ng/L		02/15/22 07:57	02/17/22 01:56	1
Perfluoropentanoic acid	1.9	U	1.9	ng/L		02/15/22 07:57	02/17/22 01:56	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	115		17 - 200	02/15/22 07:57	02/17/22 01:56	1
M2-8:2 FTS	130		33 - 200	02/15/22 07:57	02/17/22 01:56	1
13C4 PFBA	118		42 - 165	02/15/22 07:57	02/17/22 01:56	1
13C5 PFPeA	112		38 - 187	02/15/22 07:57	02/17/22 01:56	1
13C8 PFOS	121		51 - 159	02/15/22 07:57	02/17/22 01:56	1
13C8 FOSA	84		10 - 168	02/15/22 07:57	02/17/22 01:56	1
13C3 PFHxS	123		28 - 188	02/15/22 07:57	02/17/22 01:56	1

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
NEtFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
NMeFOSAA	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/10/22 17:14	02/14/22 15:41	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	78		70 - 130	02/10/22 17:14	02/14/22 15:41	1
13C2 PFDA	77		70 - 130	02/10/22 17:14	02/14/22 15:41	1
13C2 PFHxA	85		70 - 130	02/10/22 17:14	02/14/22 15:41	1

Surrogate Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		d5NEFOS (70-130)	PFDA (70-130)	PFHxA (70-130)
410-71774-1	GAC Influent	79	104	112
410-71774-1 - DL	GAC Influent	86	90	108
410-71774-2	GAC Midfluent	84	80	94
410-71774-3	GAC Effluent	85	84	95
410-71774-4	PV-2_25	85	81	95
410-71774-5	FTB01-220203	83	82	91
410-71774-6	LTB01-220203	78	77	85
LCS 410-222886/2-A	Lab Control Sample	87	88	97
LCSD 410-222886/3-A	Lab Control Sample Dup	85	86	98
MB 410-222886/1-A	Method Blank	85	85	97

Surrogate Legend

d5NEFOS = d5-NEtFOSAA

PFDA = 13C2 PFDA

PFHxA = 13C2 PFHxA

Isotope Dilution Summary

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
 SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Isotope Dilution Recovery (Acceptance Limits)						
		M262FTS (17-200)	M282FTS (33-200)	PFBA (42-165)	PFPeA (38-187)	C8PFOS (51-159)	PFOSA (10-168)	C3PFHS (28-188)
410-71774-1	GAC Influent	89	105	108	111	104	64	113
410-71774-2	GAC Midfluent	112	111	112	119	115	85	110
410-71774-3	GAC Effluent	114	112	114	108	113	82	100
410-71774-4	PV-2_25	110	112	115	110	110	76	98
410-71774-5	FTB01-220203	104	116	116	111	116	83	99
410-71774-6	LTB01-220203	115	130	118	112	121	84	123
LCS 410-223897/3-A	Lab Control Sample	107	119	113	105	109	86	111
LCSD 410-223897/4-A	Lab Control Sample Dup	116	126	117	110	117	89	108
MB 410-223897/1-A	Method Blank	109	104	112	108	109	80	102

Surrogate Legend

- M262FTS = M2-6:2 FTS
- M282FTS = M2-8:2 FTS
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C8PFOS = 13C8 PFOS
- PFOSA = 13C8 FOSA
- C3PFHS = 13C3 PFHxS

QC Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Lab Sample ID: MB 410-223897/1-A
Matrix: Water
Analysis Batch: 224540

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 223897

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	5.0	U	5.0	ng/L		02/15/22 07:57	02/16/22 23:10	1
8:2 Fluorotelomer sulfonic acid	3.0	U	3.0	ng/L		02/15/22 07:57	02/16/22 23:10	1
Perfluorobutanoic acid	5.0	U	5.0	ng/L		02/15/22 07:57	02/16/22 23:10	1
Perfluorodecanesulfonic acid	2.0	U	2.0	ng/L		02/15/22 07:57	02/16/22 23:10	1
Perfluoroheptanesulfonic acid	2.0	U	2.0	ng/L		02/15/22 07:57	02/16/22 23:10	1
Perfluorooctanesulfonamide	2.0	U	2.0	ng/L		02/15/22 07:57	02/16/22 23:10	1
Perfluoropentanoic acid	2.0	U	2.0	ng/L		02/15/22 07:57	02/16/22 23:10	1

Isotope Dilution	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	109		17 - 200	02/15/22 07:57	02/16/22 23:10	1
M2-8:2 FTS	104		33 - 200	02/15/22 07:57	02/16/22 23:10	1
13C4 PFBA	112		42 - 165	02/15/22 07:57	02/16/22 23:10	1
13C5 PFPeA	108		38 - 187	02/15/22 07:57	02/16/22 23:10	1
13C8 PFOS	109		51 - 159	02/15/22 07:57	02/16/22 23:10	1
13C8 FOSA	80		10 - 168	02/15/22 07:57	02/16/22 23:10	1
13C3 PFHxS	102		28 - 188	02/15/22 07:57	02/16/22 23:10	1

Lab Sample ID: LCS 410-223897/3-A
Matrix: Water
Analysis Batch: 224540

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 223897

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
6:2 Fluorotelomer sulfonic acid	24.3	19.9		ng/L		82	28 - 173
8:2 Fluorotelomer sulfonic acid	24.5	16.8		ng/L		68	55 - 138
Perfluorobutanoic acid	25.6	17.8		ng/L		70	59 - 136
Perfluorodecanesulfonic acid	24.7	17.4		ng/L		70	55 - 137
Perfluoroheptanesulfonic acid	24.4	19.5		ng/L		80	56 - 140
Perfluorooctanesulfonamide	25.6	25.1		ng/L		98	43 - 167
Perfluoropentanoic acid	25.6	20.1		ng/L		79	57 - 141

Isotope Dilution	LCS %Recovery	LCS Qualifier	Limits
M2-6:2 FTS	107		17 - 200
M2-8:2 FTS	119		33 - 200
13C4 PFBA	113		42 - 165
13C5 PFPeA	105		38 - 187
13C8 PFOS	109		51 - 159
13C8 FOSA	86		10 - 168
13C3 PFHxS	111		28 - 188

Lab Sample ID: LCSD 410-223897/4-A
Matrix: Water
Analysis Batch: 224540

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 223897

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
6:2 Fluorotelomer sulfonic acid	24.3	18.7		ng/L		77	28 - 173	6	30
8:2 Fluorotelomer sulfonic acid	24.5	15.9		ng/L		65	55 - 138	6	30
Perfluorobutanoic acid	25.6	19.0		ng/L		74	59 - 136	7	30
Perfluorodecanesulfonic acid	24.7	17.3		ng/L		70	55 - 137	1	30
Perfluoroheptanesulfonic acid	24.4	18.6		ng/L		76	56 - 140	5	30

QC Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified (Continued)

Lab Sample ID: LCSD 410-223897/4-A
Matrix: Water
Analysis Batch: 224540

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 223897

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Perfluorooctanesulfonamide	25.6	27.1		ng/L		106	43 - 167	8	30
Perfluoropentanoic acid	25.6	20.2		ng/L		79	57 - 141	0	30

Isotope Dilution	LCSD LCSD		Limits
	%Recovery	Qualifier	
M2-6:2 FTS	116		17 - 200
M2-8:2 FTS	126		33 - 200
13C4 PFBA	117		42 - 165
13C5 PFPeA	110		38 - 187
13C8 PFOS	117		51 - 159
13C8 FOSA	89		10 - 168
13C3 PFHxS	108		28 - 188

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MB 410-222886/1-A
Matrix: Water
Analysis Batch: 223587

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 222886

Analyte	MB MB		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Perfluorohexanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluoroheptanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorooctanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorononanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorodecanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorotridecanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorotetradecanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorobutanesulfonic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorohexanesulfonic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorooctanesulfonic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
NEtFOSAA	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
NMeFOSAA	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluoroundecanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1
Perfluorododecanoic acid	2.0	U	2.0	ng/L		02/10/22 17:14	02/14/22 12:46	1

Surrogate	MB MB		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
d5-NEtFOSAA	85		70 - 130	02/10/22 17:14	02/14/22 12:46	1
13C2 PFDA	85		70 - 130	02/10/22 17:14	02/14/22 12:46	1
13C2 PFHxA	97		70 - 130	02/10/22 17:14	02/14/22 12:46	1

Lab Sample ID: LCS 410-222886/2-A
Matrix: Water
Analysis Batch: 223587

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 222886

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Perfluorohexanoic acid	20.5	19.7		ng/L		96	70 - 130
Perfluoroheptanoic acid	20.5	19.7		ng/L		96	70 - 130
Perfluorooctanoic acid	20.5	19.6		ng/L		96	70 - 130
Perfluorononanoic acid	20.5	18.8		ng/L		92	70 - 130
Perfluorodecanoic acid	20.5	18.2		ng/L		89	70 - 130

Eurofins Lancaster Laboratories Env, LLC

QC Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: LCS 410-222886/2-A

Matrix: Water

Analysis Batch: 223587

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 222886

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec.	
							Limits	
Perfluorotridecanoic acid	20.5	16.7		ng/L		82	70 - 130	
Perfluorotetradecanoic acid	20.5	16.3		ng/L		79	70 - 130	
Perfluorobutanesulfonic acid	18.1	18.0		ng/L		99	70 - 130	
Perfluorohexanesulfonic acid	18.7	18.2		ng/L		98	70 - 130	
Perfluorooctanesulfonic acid	19.0	16.9		ng/L		89	70 - 130	
NEtFOSAA	20.5	17.2		ng/L		84	70 - 130	
NMeFOSAA	20.5	17.3		ng/L		84	70 - 130	
Perfluoroundecanoic acid	20.5	17.8		ng/L		87	70 - 130	
Perfluorododecanoic acid	20.5	17.2		ng/L		84	70 - 130	

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	87		70 - 130
13C2 PFDA	88		70 - 130
13C2 PFHxA	97		70 - 130

Lab Sample ID: LCSD 410-222886/3-A

Matrix: Water

Analysis Batch: 223587

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 222886

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec.		RPD	
							Limits		RPD	Limit
Perfluorohexanoic acid	20.5	20.0		ng/L		98	70 - 130		1	30
Perfluoroheptanoic acid	20.5	20.2		ng/L		99	70 - 130		3	30
Perfluorooctanoic acid	20.5	19.7		ng/L		96	70 - 130		1	30
Perfluorononanoic acid	20.5	19.2		ng/L		94	70 - 130		2	30
Perfluorodecanoic acid	20.5	17.8		ng/L		87	70 - 130		2	30
Perfluorotridecanoic acid	20.5	16.5		ng/L		80	70 - 130		1	30
Perfluorotetradecanoic acid	20.5	16.2		ng/L		79	70 - 130		0	30
Perfluorobutanesulfonic acid	18.1	17.9		ng/L		99	70 - 130		0	30
Perfluorohexanesulfonic acid	18.7	18.2		ng/L		97	70 - 130		0	30
Perfluorooctanesulfonic acid	19.0	17.0		ng/L		90	70 - 130		0	30
NEtFOSAA	20.5	17.2		ng/L		84	70 - 130		0	30
NMeFOSAA	20.5	17.3		ng/L		84	70 - 130		0	30
Perfluoroundecanoic acid	20.5	17.4		ng/L		85	70 - 130		2	30
Perfluorododecanoic acid	20.5	17.9		ng/L		87	70 - 130		4	30

Surrogate	LCSD LCSD		Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	85		70 - 130
13C2 PFDA	86		70 - 130
13C2 PFHxA	98		70 - 130

QC Association Summary

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
 SDG: HOO

LCMS

Prep Batch: 222886

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-71774-1 - DL	GAC Influent	Total/NA	Water	537 DW	
410-71774-1	GAC Influent	Total/NA	Water	537 DW	
410-71774-2	GAC Midfluent	Total/NA	Water	537 DW	
410-71774-3	GAC Effluent	Total/NA	Water	537 DW	
410-71774-4	PV-2_25	Total/NA	Water	537 DW	
410-71774-5	FTB01-220203	Total/NA	Water	537 DW	
410-71774-6	LTB01-220203	Total/NA	Water	537 DW	
MB 410-222886/1-A	Method Blank	Total/NA	Water	537 DW	
LCS 410-222886/2-A	Lab Control Sample	Total/NA	Water	537 DW	
LCSD 410-222886/3-A	Lab Control Sample Dup	Total/NA	Water	537 DW	

Analysis Batch: 223587

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-71774-1	GAC Influent	Total/NA	Water	537 DW	222886
410-71774-2	GAC Midfluent	Total/NA	Water	537 DW	222886
410-71774-3	GAC Effluent	Total/NA	Water	537 DW	222886
410-71774-4	PV-2_25	Total/NA	Water	537 DW	222886
410-71774-5	FTB01-220203	Total/NA	Water	537 DW	222886
410-71774-6	LTB01-220203	Total/NA	Water	537 DW	222886
MB 410-222886/1-A	Method Blank	Total/NA	Water	537 DW	222886
LCS 410-222886/2-A	Lab Control Sample	Total/NA	Water	537 DW	222886
LCSD 410-222886/3-A	Lab Control Sample Dup	Total/NA	Water	537 DW	222886

Prep Batch: 223897

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-71774-1	GAC Influent	Total/NA	Water	537 (Mod)	
410-71774-2	GAC Midfluent	Total/NA	Water	537 (Mod)	
410-71774-3	GAC Effluent	Total/NA	Water	537 (Mod)	
410-71774-4	PV-2_25	Total/NA	Water	537 (Mod)	
410-71774-5	FTB01-220203	Total/NA	Water	537 (Mod)	
410-71774-6	LTB01-220203	Total/NA	Water	537 (Mod)	
MB 410-223897/1-A	Method Blank	Total/NA	Water	537 (Mod)	
LCS 410-223897/3-A	Lab Control Sample	Total/NA	Water	537 (Mod)	
LCSD 410-223897/4-A	Lab Control Sample Dup	Total/NA	Water	537 (Mod)	

Analysis Batch: 223901

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-71774-1 - DL	GAC Influent	Total/NA	Water	537 DW	222886

Analysis Batch: 224540

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-71774-1	GAC Influent	Total/NA	Water	537 (Mod)	223897
410-71774-2	GAC Midfluent	Total/NA	Water	537 (Mod)	223897
410-71774-3	GAC Effluent	Total/NA	Water	537 (Mod)	223897
410-71774-4	PV-2_25	Total/NA	Water	537 (Mod)	223897
410-71774-5	FTB01-220203	Total/NA	Water	537 (Mod)	223897
410-71774-6	LTB01-220203	Total/NA	Water	537 (Mod)	223897
MB 410-223897/1-A	Method Blank	Total/NA	Water	537 (Mod)	223897
LCS 410-223897/3-A	Lab Control Sample	Total/NA	Water	537 (Mod)	223897
LCSD 410-223897/4-A	Lab Control Sample Dup	Total/NA	Water	537 (Mod)	223897

Lab Chronicle

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: GAC Influent

Lab Sample ID: 410-71774-1

Date Collected: 02/03/22 14:40

Matrix: Water

Date Received: 02/04/22 10:01

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537 (Mod)			223897	02/15/22 07:57	RC3V	ELLE
Total/NA	Analysis	537 (Mod)		1	224540	02/17/22 00:50	QD9Y	ELLE
Total/NA	Prep	537 DW			222886	02/10/22 17:14	GU2F	ELLE
Total/NA	Analysis	537 DW		1	223587	02/14/22 15:53	PY4D	ELLE
Total/NA	Prep	537 DW	DL		222886	02/10/22 17:14	GU2F	ELLE
Total/NA	Analysis	537 DW	DL	10	223901	02/15/22 09:53	DCS9	ELLE

Client Sample ID: GAC Midfluent

Lab Sample ID: 410-71774-2

Date Collected: 02/03/22 14:45

Matrix: Water

Date Received: 02/04/22 10:01

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537 (Mod)			223897	02/15/22 07:57	RC3V	ELLE
Total/NA	Analysis	537 (Mod)		1	224540	02/17/22 01:01	QD9Y	ELLE
Total/NA	Prep	537 DW			222886	02/10/22 17:14	GU2F	ELLE
Total/NA	Analysis	537 DW		1	223587	02/14/22 14:55	PY4D	ELLE

Client Sample ID: GAC Effluent

Lab Sample ID: 410-71774-3

Date Collected: 02/03/22 14:48

Matrix: Water

Date Received: 02/04/22 10:01

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537 (Mod)			223897	02/15/22 07:57	RC3V	ELLE
Total/NA	Analysis	537 (Mod)		1	224540	02/17/22 01:12	QD9Y	ELLE
Total/NA	Prep	537 DW			222886	02/10/22 17:14	GU2F	ELLE
Total/NA	Analysis	537 DW		1	223587	02/14/22 15:06	PY4D	ELLE

Client Sample ID: PV-2_25

Lab Sample ID: 410-71774-4

Date Collected: 02/03/22 14:50

Matrix: Water

Date Received: 02/04/22 10:01

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537 (Mod)			223897	02/15/22 07:57	RC3V	ELLE
Total/NA	Analysis	537 (Mod)		1	224540	02/17/22 01:23	QD9Y	ELLE
Total/NA	Prep	537 DW			222886	02/10/22 17:14	GU2F	ELLE
Total/NA	Analysis	537 DW		1	223587	02/14/22 15:18	PY4D	ELLE

Client Sample ID: FTB01-220203

Lab Sample ID: 410-71774-5

Date Collected: 02/03/22 14:55

Matrix: Water

Date Received: 02/04/22 10:01

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537 (Mod)			223897	02/15/22 07:57	RC3V	ELLE
Total/NA	Analysis	537 (Mod)		1	224540	02/17/22 01:34	QD9Y	ELLE
Total/NA	Prep	537 DW			222886	02/10/22 17:14	GU2F	ELLE
Total/NA	Analysis	537 DW		1	223587	02/14/22 15:29	PY4D	ELLE

Lab Chronicle

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Client Sample ID: LTB01-220203

Lab Sample ID: 410-71774-6

Date Collected: 02/03/22 00:00

Matrix: Water

Date Received: 02/04/22 10:01

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537 (Mod)			223897	02/15/22 07:57	RC3V	ELLE
Total/NA	Analysis	537 (Mod)		1	224540	02/17/22 01:56	QD9Y	ELLE
Total/NA	Prep	537 DW			222886	02/10/22 17:14	GU2F	ELLE
Total/NA	Analysis	537 DW		1	223587	02/14/22 15:41	PY4D	ELLE

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Env, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300



Accreditation/Certification Summary

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
 SDG: HOO

Laboratory: Eurofins Lancaster Laboratories Env, LLC

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
New York	NELAP	10670	04-01-22

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
537 (Mod)	537 (Mod)	Water	6:2 Fluorotelomer sulfonic acid
537 (Mod)	537 (Mod)	Water	8:2 Fluorotelomer sulfonic acid
537 (Mod)	537 (Mod)	Water	Perfluorobutanoic acid
537 (Mod)	537 (Mod)	Water	Perfluorodecanesulfonic acid
537 (Mod)	537 (Mod)	Water	Perfluoroheptanesulfonic acid
537 (Mod)	537 (Mod)	Water	Perfluorooctanesulfonamide
537 (Mod)	537 (Mod)	Water	Perfluoropentanoic acid
537 DW	537 DW	Water	NEtFOSAA
537 DW	537 DW	Water	NMeFOSAA
537 DW	537 DW	Water	Perfluorobutanesulfonic acid
537 DW	537 DW	Water	Perfluorodecanoic acid
537 DW	537 DW	Water	Perfluorododecanoic acid
537 DW	537 DW	Water	Perfluoroheptanoic acid
537 DW	537 DW	Water	Perfluorohexanesulfonic acid
537 DW	537 DW	Water	Perfluorohexanoic acid
537 DW	537 DW	Water	Perfluorononanoic acid
537 DW	537 DW	Water	Perfluorooctanesulfonic acid
537 DW	537 DW	Water	Perfluorooctanoic acid
537 DW	537 DW	Water	Perfluorotetradecanoic acid
537 DW	537 DW	Water	Perfluorotridecanoic acid
537 DW	537 DW	Water	Perfluoroundecanoic acid



Method Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Method	Method Description	Protocol	Laboratory
537 (Mod)	EPA 537 Version 1.1 modified	EPA	ELLE
537 DW	Perfluorinated Alkyl Acids (LC/MS)	EPA	ELLE
537 (Mod)	537 Version 1.1 modified	EPA	ELLE
537 DW	Extraction of Perfluorinated Alkyl Acids	EPA	ELLE

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Env, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300



Sample Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-71774-1
SDG: HOO

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
410-71774-1	GAC Influent	Water	02/03/22 14:40	02/04/22 10:01
410-71774-2	GAC Midfluent	Water	02/03/22 14:45	02/04/22 10:01
410-71774-3	GAC Effluent	Water	02/03/22 14:48	02/04/22 10:01
410-71774-4	PV-2_25	Water	02/03/22 14:50	02/04/22 10:01
410-71774-5	FTB01-220203	Water	02/03/22 14:55	02/04/22 10:01
410-71774-6	LTB01-220203	Water	02/03/22 00:00	02/04/22 10:01

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Login Sample Receipt Checklist

Client: CT Male Associates DPC

Job Number: 410-71774-1

SDG Number: HOO

Login Number: 71774

List Source: Eurofins Lancaster Laboratories Env, LLC

List Number: 1

Creator: Reiff, Nicole L

Question	Answer	Comment
The cooler's custody seal is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable ($\leq 6^{\circ}\text{C}$, not frozen).	True	
Cooler Temperature is recorded.	True	
WV: Container Temperature is acceptable ($\leq 6^{\circ}\text{C}$, not frozen).	N/A	
WV: Container Temperature is recorded.	N/A	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses.	True	
Is the Field Sampler's name present on COC?	True	
Sample custody seals are intact.	True	

