



ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Kirk Moline
CT Male Associates DPC
50 Century Hill Dr
Latham, New York 12110

Generated 10/11/2023 9:57:52 PM

JOB DESCRIPTION

Hoosick Falls WTP
SDG NUMBER HOO

JOB NUMBER


410-145941-1

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

Analytical test results meet all requirements of the associated regulatory program (i.e., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis.

Authorization



Generated
10/11/2023 9:57:52 PM

Authorized for release by
Kelly Tessier, Project Manager
kelly.tessier@et.eurofinsus.com
(717)205-7820

Compliance Statement

Analytical test results meet all requirements of the associated regulatory program (e.g., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis. Data qualifiers are applied to note exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- QC results that exceed the upper limits and are associated with non-detect samples are qualified but further narration is not required since the bias is high and does not change a non-detect result. Further narration is also not required with QC blank detection when the associated sample concentration is non-detect or more than ten times the level in the blank.
- Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD is performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Measurement uncertainty values, as applicable, are available upon request.

Test results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" and tested in the laboratory are not performed within 15 minutes of collection.

This report shall not be reproduced except in full, without the written approval of the laboratory.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. The foregoing express warranty is exclusive and is given in lieu of all other warranties, expressed or implied, except as otherwise agreed. We disclaim any other warranties, expressed or implied, including a warranty of fitness for particular purpose and warranty of merchantability. In no event shall Eurofins Lancaster Laboratories Environmental, LLC be liable for indirect, special, consequential, or incidental damages including, but not limited to, damages for loss of profit or goodwill regardless of (A) the negligence (either sole or concurrent) of Eurofins Lancaster Laboratories Environmental and (B) whether Eurofins Lancaster Laboratories Environmental has been informed of the possibility of such damages. We accept no legal responsibility for the purposes for which the client uses the test results. Except as otherwise agreed, no purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.





Table of Contents

Cover Page	1
Table of Contents	4
Definitions/Glossary	5
Case Narrative	6
Detection Summary	7
Client Sample Results	8
Surrogate Summary	13
Isotope Dilution Summary	14
QC Sample Results	15
QC Association Summary	18
Lab Chronicle	19
Certification Summary	21
Method Summary	22
Sample Summary	23
Chain of Custody	24
Receipt Checklists	25

Definitions/Glossary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Qualifiers

LCMS

Qualifier	Qualifier Description
U	Indicates the analyte was analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Job ID: 410-145941-1

Laboratory: Eurofins Lancaster Laboratories Environment Testing, LLC

Narrative

**Job Narrative
410-145941-1**

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers are applied to indicate exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method. Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 10/6/2023 9:45 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 1.5°C

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.



Detection Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Client Sample ID: GAC INFLUENT

Lab Sample ID: 410-145941-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	3.6		1.8	ng/L	1		537 (Mod)	Total/NA
Perfluoropentanoic acid	3.4		1.8	ng/L	1		537 (Mod)	Total/NA
Perfluoroheptanoic acid	8.9		1.9	ng/L	1		EPA 537.1	Total/NA
Perfluorohexanoic acid	8.5		1.9	ng/L	1		EPA 537.1	Total/NA
Perfluorooctanesulfonic acid	3.7		1.9	ng/L	1		EPA 537.1	Total/NA
Perfluorooctanoic acid - DL	260		19	ng/L	10		EPA 537.1	Total/NA

Client Sample ID: GAC MIDFLUENT

Lab Sample ID: 410-145941-2

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	2.5		1.7	ng/L	1		537 (Mod)	Total/NA

Client Sample ID: GAC EFFLUENT

Lab Sample ID: 410-145941-3

No Detections.

Client Sample ID: FTB01-231005

Lab Sample ID: 410-145941-4

No Detections.

Client Sample ID: LTB01-231005

Lab Sample ID: 410-145941-5

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Environment Testing, LLC

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Client Sample ID: GAC INFLUENT

Lab Sample ID: 410-145941-1

Date Collected: 10/05/23 09:30

Matrix: Water

Date Received: 10/06/23 09:45

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 17:44	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 17:44	1
Perfluorobutanoic acid	3.6		1.8	ng/L		10/09/23 07:26	10/10/23 17:44	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 17:44	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 17:44	1
Perfluorooctanesulfonamide	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 17:44	1
Perfluoropentanoic acid	3.4		1.8	ng/L		10/09/23 07:26	10/10/23 17:44	1
<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
M2-6:2 FTS	103		40 - 200			10/09/23 07:26	10/10/23 17:44	1
M2-8:2 FTS	117		37 - 200			10/09/23 07:26	10/10/23 17:44	1
13C4 PFBA	89		22 - 174			10/09/23 07:26	10/10/23 17:44	1
13C5 PFPeA	97		33 - 196			10/09/23 07:26	10/10/23 17:44	1
13C8 PFOS	97		59 - 155			10/09/23 07:26	10/10/23 17:44	1
13C8 FOSA	64		10 - 155			10/09/23 07:26	10/10/23 17:44	1
13C3 PFHxS	105		48 - 169			10/09/23 07:26	10/10/23 17:44	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
NMeFOSAA	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorobutanesulfonic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorodecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorododecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluoroheptanoic acid	8.9		1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorohexanesulfonic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorohexanoic acid	8.5		1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorononanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorooctanesulfonic acid	3.7		1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorotetradecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluorotridecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
Perfluoroundecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:36	1
<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C2 PFDA	108		70 - 130			10/09/23 16:01	10/10/23 18:36	1
13C2 PFHxA	101		70 - 130			10/09/23 16:01	10/10/23 18:36	1
d5-NEtFOSAA	103		70 - 130			10/09/23 16:01	10/10/23 18:36	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018 - DL

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanoic acid	260		19	ng/L		10/09/23 16:01	10/11/23 19:23	10
<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C2 PFDA	71		70 - 130			10/09/23 16:01	10/11/23 19:23	10
13C2 PFHxA	71		70 - 130			10/09/23 16:01	10/11/23 19:23	10
d5-NEtFOSAA	82		70 - 130			10/09/23 16:01	10/11/23 19:23	10

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Client Sample ID: GAC MIDFLUENT

Lab Sample ID: 410-145941-2

Date Collected: 10/05/23 09:35

Matrix: Water

Date Received: 10/06/23 09:45

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:08	1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:08	1
Perfluorobutanoic acid	2.5		1.7	ng/L		10/09/23 07:26	10/10/23 18:08	1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:08	1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:08	1
Perfluorooctanesulfonamide	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:08	1
Perfluoropentanoic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:08	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	115		40 - 200	10/09/23 07:26	10/10/23 18:08	1
M2-8:2 FTS	117		37 - 200	10/09/23 07:26	10/10/23 18:08	1
13C4 PFBA	85		22 - 174	10/09/23 07:26	10/10/23 18:08	1
13C5 PFPeA	90		33 - 196	10/09/23 07:26	10/10/23 18:08	1
13C8 PFOS	102		59 - 155	10/09/23 07:26	10/10/23 18:08	1
13C8 FOSA	81		10 - 155	10/09/23 07:26	10/10/23 18:08	1
13C3 PFHxS	104		48 - 169	10/09/23 07:26	10/10/23 18:08	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
NMeFOSAA	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorobutanesulfonic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorodecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorododecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluoroheptanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorohexanesulfonic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorohexanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorononanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorooctanesulfonic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorooctanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorotetradecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluorotridecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1
Perfluoroundecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 18:48	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	100		70 - 130	10/09/23 16:01	10/10/23 18:48	1
13C2 PFHxA	99		70 - 130	10/09/23 16:01	10/10/23 18:48	1
d5-NEtFOSAA	101		70 - 130	10/09/23 16:01	10/10/23 18:48	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Client Sample ID: GAC EFFLUENT

Lab Sample ID: 410-145941-3

Date Collected: 10/05/23 09:40

Matrix: Water

Date Received: 10/06/23 09:45

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:21	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:21	1
Perfluorobutanoic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:21	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:21	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:21	1
Perfluorooctanesulfonamide	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:21	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:21	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	104		40 - 200	10/09/23 07:26	10/10/23 18:21	1
M2-8:2 FTS	100		37 - 200	10/09/23 07:26	10/10/23 18:21	1
13C4 PFBA	86		22 - 174	10/09/23 07:26	10/10/23 18:21	1
13C5 PFPeA	88		33 - 196	10/09/23 07:26	10/10/23 18:21	1
13C8 PFOS	98		59 - 155	10/09/23 07:26	10/10/23 18:21	1
13C8 FOSA	79		10 - 155	10/09/23 07:26	10/10/23 18:21	1
13C3 PFHxS	94		48 - 169	10/09/23 07:26	10/10/23 18:21	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
NMeFOSAA	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorohexanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 18:59	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	101		70 - 130	10/09/23 16:01	10/10/23 18:59	1
13C2 PFHxA	105		70 - 130	10/09/23 16:01	10/10/23 18:59	1
d5-NEtFOSAA	107		70 - 130	10/09/23 16:01	10/10/23 18:59	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Client Sample ID: FTB01-231005

Lab Sample ID: 410-145941-4

Date Collected: 10/05/23 09:45

Matrix: Water

Date Received: 10/06/23 09:45

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:34	1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:34	1
Perfluorobutanoic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:34	1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:34	1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:34	1
Perfluorooctanesulfonamide	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:34	1
Perfluoropentanoic acid	1.7	U	1.7	ng/L		10/09/23 07:26	10/10/23 18:34	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	103		40 - 200	10/09/23 07:26	10/10/23 18:34	1
M2-8:2 FTS	102		37 - 200	10/09/23 07:26	10/10/23 18:34	1
13C4 PFBA	85		22 - 174	10/09/23 07:26	10/10/23 18:34	1
13C5 PFPeA	96		33 - 196	10/09/23 07:26	10/10/23 18:34	1
13C8 PFOS	97		59 - 155	10/09/23 07:26	10/10/23 18:34	1
13C8 FOSA	81		10 - 155	10/09/23 07:26	10/10/23 18:34	1
13C3 PFHxS	93		48 - 169	10/09/23 07:26	10/10/23 18:34	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
NMeFOSAA	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorobutanesulfonic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorodecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorododecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluoroheptanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorohexanesulfonic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorohexanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorononanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorooctanesulfonic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorooctanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorotetradecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluorotridecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1
Perfluoroundecanoic acid	1.9	U	1.9	ng/L		10/09/23 16:01	10/10/23 19:11	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	103		70 - 130	10/09/23 16:01	10/10/23 19:11	1
13C2 PFHxA	97		70 - 130	10/09/23 16:01	10/10/23 19:11	1
d5-NEtFOSAA	109		70 - 130	10/09/23 16:01	10/10/23 19:11	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Client Sample ID: LTB01-231005

Lab Sample ID: 410-145941-5

Date Collected: 10/05/23 00:00

Matrix: Water

Date Received: 10/06/23 09:45

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:46	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:46	1
Perfluorobutanoic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:46	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:46	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:46	1
Perfluorooctanesulfonamide	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:46	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		10/09/23 07:26	10/10/23 18:46	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	106		40 - 200	10/09/23 07:26	10/10/23 18:46	1
M2-8:2 FTS	101		37 - 200	10/09/23 07:26	10/10/23 18:46	1
13C4 PFBA	93		22 - 174	10/09/23 07:26	10/10/23 18:46	1
13C5 PFPeA	94		33 - 196	10/09/23 07:26	10/10/23 18:46	1
13C8 PFOS	97		59 - 155	10/09/23 07:26	10/10/23 18:46	1
13C8 FOSA	82		10 - 155	10/09/23 07:26	10/10/23 18:46	1
13C3 PFHxS	91		48 - 169	10/09/23 07:26	10/10/23 18:46	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
NMeFOSAA	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorohexanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		10/09/23 16:01	10/10/23 19:34	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	99		70 - 130	10/09/23 16:01	10/10/23 19:34	1
13C2 PFHxA	99		70 - 130	10/09/23 16:01	10/10/23 19:34	1
d5-NEtFOSAA	110		70 - 130	10/09/23 16:01	10/10/23 19:34	1

Surrogate Summary

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
 SDG: HOO

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		PFDA (70-130)	PFHxA (70-130)	d5NEFOS (70-130)
410-145941-1	GAC INFLUENT	108	101	103
410-145941-1 - DL	GAC INFLUENT	71	71	82
410-145941-2	GAC MIDFLUENT	100	99	101
410-145941-3	GAC EFFLUENT	101	105	107
410-145941-4	FTB01-231005	103	97	109
410-145941-5	LTB01-231005	99	99	110
LCS 410-429105/2-A	Lab Control Sample	105	100	96
MB 410-429105/1-A	Method Blank	100	94	99

Surrogate Legend

PFDA = 13C2 PFDA

PFHxA = 13C2 PFHxA

d5NEFOS = d5-NEtFOSAA



Isotope Dilution Summary

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
 SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Isotope Dilution Recovery (Acceptance Limits)						
		M262FTS (40-200)	M282FTS (37-200)	PFBA (22-174)	PFPeA (33-196)	C8PFOS (59-155)	PFOSA (10-155)	C3PFHS (48-169)
410-145941-1	GAC INFLUENT	103	117	89	97	97	64	105
410-145941-2	GAC MIDFLUENT	115	117	85	90	102	81	104
410-145941-3	GAC EFFLUENT	104	100	86	88	98	79	94
410-145941-4	FTB01-231005	103	102	85	96	97	81	93
410-145941-5	LTB01-231005	106	101	93	94	97	82	91
LCS 410-428825/2-A	Lab Control Sample	98	105	71	94	100	93	98
LCSD 410-428825/3-A	Lab Control Sample Dup	101	103	91	99	104	88	100
MB 410-428825/1-A	Method Blank	104	103	93	100	100	80	100

Surrogate Legend

- M262FTS = M2-6:2 FTS
- M282FTS = M2-8:2 FTS
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C8PFOS = 13C8 PFOS
- PFOSA = 13C8 FOSA
- C3PFHS = 13C3 PFHxS

QC Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Lab Sample ID: MB 410-428825/1-A
Matrix: Water
Analysis Batch: 429483

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 428825

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
6:2 Fluorotelomer sulfonic acid	2.0	U	2.0	ng/L		10/09/23 07:26	10/10/23 14:42	1
8:2 Fluorotelomer sulfonic acid	2.0	U	2.0	ng/L		10/09/23 07:26	10/10/23 14:42	1
Perfluorobutanoic acid	2.0	U	2.0	ng/L		10/09/23 07:26	10/10/23 14:42	1
Perfluorodecanesulfonic acid	2.0	U	2.0	ng/L		10/09/23 07:26	10/10/23 14:42	1
Perfluoroheptanesulfonic acid	2.0	U	2.0	ng/L		10/09/23 07:26	10/10/23 14:42	1
Perfluorooctanesulfonamide	2.0	U	2.0	ng/L		10/09/23 07:26	10/10/23 14:42	1
Perfluoropentanoic acid	2.0	U	2.0	ng/L		10/09/23 07:26	10/10/23 14:42	1

Isotope Dilution	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
M2-6:2 FTS	104		40 - 200	10/09/23 07:26	10/10/23 14:42	1
M2-8:2 FTS	103		37 - 200	10/09/23 07:26	10/10/23 14:42	1
13C4 PFBA	93		22 - 174	10/09/23 07:26	10/10/23 14:42	1
13C5 PFPeA	100		33 - 196	10/09/23 07:26	10/10/23 14:42	1
13C8 PFOS	100		59 - 155	10/09/23 07:26	10/10/23 14:42	1
13C8 FOSA	80		10 - 155	10/09/23 07:26	10/10/23 14:42	1
13C3 PFHxS	100		48 - 169	10/09/23 07:26	10/10/23 14:42	1

Lab Sample ID: LCS 410-428825/2-A
Matrix: Water
Analysis Batch: 429483

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 428825

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
8:2 Fluorotelomer sulfonic acid	24.5	24.8		ng/L		101	55 - 134
Perfluorobutanoic acid	25.6	24.1		ng/L		94	58 - 130
Perfluorodecanesulfonic acid	24.7	23.8		ng/L		97	55 - 130
Perfluoroheptanesulfonic acid	24.4	23.1		ng/L		95	59 - 130
Perfluorooctanesulfonamide	25.6	26.7		ng/L		104	67 - 132
Perfluoropentanoic acid	25.6	25.1		ng/L		98	60 - 130

Isotope Dilution	LCS	LCS	Limits
	%Recovery	Qualifier	
M2-6:2 FTS	98		40 - 200
M2-8:2 FTS	105		37 - 200
13C4 PFBA	71		22 - 174
13C5 PFPeA	94		33 - 196
13C8 PFOS	100		59 - 155
13C8 FOSA	93		10 - 155
13C3 PFHxS	98		48 - 169

Lab Sample ID: LCSD 410-428825/3-A
Matrix: Water
Analysis Batch: 429483

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 428825

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	Limit
8:2 Fluorotelomer sulfonic acid	24.5	25.5		ng/L		104	55 - 134	3	30
Perfluorobutanoic acid	25.6	24.2		ng/L		95	58 - 130	1	30
Perfluorodecanesulfonic acid	24.7	23.6		ng/L		95	55 - 130	1	30
Perfluoroheptanesulfonic acid	24.4	23.0		ng/L		94	59 - 130	0	30

QC Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified (Continued)

Lab Sample ID: LCSD 410-428825/3-A
Matrix: Water
Analysis Batch: 429483

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 428825

Analyte	Spike	LCS	LCS	Unit	D	%Rec	%Rec	RPD	Limit
	Added	Result	Qualifier						
Perfluorooctanesulfonamide	25.6	26.8		ng/L		105	67 - 132	0	30
Perfluoropentanoic acid	25.6	24.6		ng/L		96	60 - 130	2	30

Isotope Dilution	LCS	LCS	Limits
	%Recovery	Qualifier	
M2-6:2 FTS	101		40 - 200
M2-8:2 FTS	103		37 - 200
13C4 PFBA	91		22 - 174
13C5 PFPeA	99		33 - 196
13C8 PFOS	104		59 - 155
13C8 FOSA	88		10 - 155
13C3 PFHxS	100		48 - 169

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Lab Sample ID: MB 410-429105/1-A
Matrix: Water
Analysis Batch: 429496

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 429105

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
NEtFOSAA	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
NMeFOSAA	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorobutanesulfonic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorodecanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorododecanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluoroheptanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorohexanesulfonic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorohexanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorononanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorooctanesulfonic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorooctanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorotetradecanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluorotridecanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1
Perfluoroundecanoic acid	2.0	U	2.0	ng/L		10/09/23 16:01	10/10/23 16:29	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C2 PFDA	100		70 - 130	10/09/23 16:01	10/10/23 16:29	1
13C2 PFHxA	94		70 - 130	10/09/23 16:01	10/10/23 16:29	1
d5-NEtFOSAA	99		70 - 130	10/09/23 16:01	10/10/23 16:29	1

Lab Sample ID: LCS 410-429105/2-A
Matrix: Water
Analysis Batch: 429496

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 429105

Analyte	Spike	LCS	LCS	Unit	D	%Rec	%Rec	Limit
	Added	Result	Qualifier					
NEtFOSAA	60.0	48.0		ng/L		80	70 - 130	
NMeFOSAA	60.0	51.5		ng/L		86	70 - 130	
Perfluorobutanesulfonic acid	53.1	53.3		ng/L		100	70 - 130	
Perfluorodecanoic acid	60.0	54.7		ng/L		91	70 - 130	
Perfluorododecanoic acid	60.0	52.8		ng/L		88	70 - 130	

QC Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018 (Continued)

Lab Sample ID: LCS 410-429105/2-A

Matrix: Water

Analysis Batch: 429496

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 429105

Analyte	Spike	LCS	LCS	Unit	D	%Rec	%Rec Limits
	Added	Result	Qualifier				
Perfluoroheptanoic acid	60.0	52.0		ng/L		87	70 - 130
Perfluorohexanesulfonic acid	54.7	53.1		ng/L		97	70 - 130
Perfluorohexanoic acid	60.0	54.9		ng/L		91	70 - 130
Perfluorononanoic acid	60.0	53.1		ng/L		89	70 - 130
Perfluorooctanesulfonic acid	55.5	51.5		ng/L		93	70 - 130
Perfluorooctanoic acid	60.0	56.0		ng/L		93	70 - 130
Perfluorotetradecanoic acid	60.0	51.8		ng/L		86	70 - 130
Perfluorotridecanoic acid	60.0	50.5		ng/L		84	70 - 130
Perfluoroundecanoic acid	60.0	50.5		ng/L		84	70 - 130

Surrogate	LCS		Limits
	%Recovery	Qualifier	
13C2 PFDA	105		70 - 130
13C2 PFHxA	100		70 - 130
d5-NEtFOSAA	96		70 - 130

QC Association Summary

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
 SDG: HOO

LCMS

Prep Batch: 428825

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-145941-1	GAC INFLUENT	Total/NA	Water	SPE	
410-145941-2	GAC MIDFLUENT	Total/NA	Water	SPE	
410-145941-3	GAC EFFLUENT	Total/NA	Water	SPE	
410-145941-4	FTB01-231005	Total/NA	Water	SPE	
410-145941-5	LTB01-231005	Total/NA	Water	SPE	
MB 410-428825/1-A	Method Blank	Total/NA	Water	SPE	
LCS 410-428825/2-A	Lab Control Sample	Total/NA	Water	SPE	
LCSD 410-428825/3-A	Lab Control Sample Dup	Total/NA	Water	SPE	

Prep Batch: 429105

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-145941-1	GAC INFLUENT	Total/NA	Water	537.1 DW Prep	
410-145941-1 - DL	GAC INFLUENT	Total/NA	Water	537.1 DW Prep	
410-145941-2	GAC MIDFLUENT	Total/NA	Water	537.1 DW Prep	
410-145941-3	GAC EFFLUENT	Total/NA	Water	537.1 DW Prep	
410-145941-4	FTB01-231005	Total/NA	Water	537.1 DW Prep	
410-145941-5	LTB01-231005	Total/NA	Water	537.1 DW Prep	
MB 410-429105/1-A	Method Blank	Total/NA	Water	537.1 DW Prep	
LCS 410-429105/2-A	Lab Control Sample	Total/NA	Water	537.1 DW Prep	

Analysis Batch: 429483

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-145941-1	GAC INFLUENT	Total/NA	Water	537 (Mod)	428825
410-145941-2	GAC MIDFLUENT	Total/NA	Water	537 (Mod)	428825
410-145941-3	GAC EFFLUENT	Total/NA	Water	537 (Mod)	428825
410-145941-4	FTB01-231005	Total/NA	Water	537 (Mod)	428825
410-145941-5	LTB01-231005	Total/NA	Water	537 (Mod)	428825
MB 410-428825/1-A	Method Blank	Total/NA	Water	537 (Mod)	428825
LCS 410-428825/2-A	Lab Control Sample	Total/NA	Water	537 (Mod)	428825
LCSD 410-428825/3-A	Lab Control Sample Dup	Total/NA	Water	537 (Mod)	428825

Analysis Batch: 429496

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-145941-1	GAC INFLUENT	Total/NA	Water	EPA 537.1	429105
410-145941-2	GAC MIDFLUENT	Total/NA	Water	EPA 537.1	429105
410-145941-3	GAC EFFLUENT	Total/NA	Water	EPA 537.1	429105
410-145941-4	FTB01-231005	Total/NA	Water	EPA 537.1	429105
410-145941-5	LTB01-231005	Total/NA	Water	EPA 537.1	429105
MB 410-429105/1-A	Method Blank	Total/NA	Water	EPA 537.1	429105
LCS 410-429105/2-A	Lab Control Sample	Total/NA	Water	EPA 537.1	429105

Analysis Batch: 430210

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-145941-1 - DL	GAC INFLUENT	Total/NA	Water	EPA 537.1	429105

Lab Chronicle

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Client Sample ID: GAC INFLUENT

Lab Sample ID: 410-145941-1

Date Collected: 10/05/23 09:30

Matrix: Water

Date Received: 10/06/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			428825	M4QQ	ELLE	10/09/23 07:26
Total/NA	Analysis	537 (Mod)		1	429483	VK3G	ELLE	10/10/23 17:44
Total/NA	Prep	537.1 DW Prep			429105	K6GF	ELLE	10/09/23 16:01
Total/NA	Analysis	EPA 537.1		1	429496	WR4P	ELLE	10/10/23 18:36
Total/NA	Prep	537.1 DW Prep	DL		429105	K6GF	ELLE	10/09/23 16:01
Total/NA	Analysis	EPA 537.1	DL	10	430210	WR4P	ELLE	10/11/23 19:23

Client Sample ID: GAC MIDFLUENT

Lab Sample ID: 410-145941-2

Date Collected: 10/05/23 09:35

Matrix: Water

Date Received: 10/06/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			428825	M4QQ	ELLE	10/09/23 07:26
Total/NA	Analysis	537 (Mod)		1	429483	VK3G	ELLE	10/10/23 18:08
Total/NA	Prep	537.1 DW Prep			429105	K6GF	ELLE	10/09/23 16:01
Total/NA	Analysis	EPA 537.1		1	429496	WR4P	ELLE	10/10/23 18:48

Client Sample ID: GAC EFFLUENT

Lab Sample ID: 410-145941-3

Date Collected: 10/05/23 09:40

Matrix: Water

Date Received: 10/06/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			428825	M4QQ	ELLE	10/09/23 07:26
Total/NA	Analysis	537 (Mod)		1	429483	VK3G	ELLE	10/10/23 18:21
Total/NA	Prep	537.1 DW Prep			429105	K6GF	ELLE	10/09/23 16:01
Total/NA	Analysis	EPA 537.1		1	429496	WR4P	ELLE	10/10/23 18:59

Client Sample ID: FTB01-231005

Lab Sample ID: 410-145941-4

Date Collected: 10/05/23 09:45

Matrix: Water

Date Received: 10/06/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			428825	M4QQ	ELLE	10/09/23 07:26
Total/NA	Analysis	537 (Mod)		1	429483	VK3G	ELLE	10/10/23 18:34
Total/NA	Prep	537.1 DW Prep			429105	K6GF	ELLE	10/09/23 16:01
Total/NA	Analysis	EPA 537.1		1	429496	WR4P	ELLE	10/10/23 19:11

Client Sample ID: LTB01-231005

Lab Sample ID: 410-145941-5

Date Collected: 10/05/23 00:00

Matrix: Water

Date Received: 10/06/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			428825	M4QQ	ELLE	10/09/23 07:26
Total/NA	Analysis	537 (Mod)		1	429483	VK3G	ELLE	10/10/23 18:46
Total/NA	Prep	537.1 DW Prep			429105	K6GF	ELLE	10/09/23 16:01
Total/NA	Analysis	EPA 537.1		1	429496	WR4P	ELLE	10/10/23 19:34

Lab Chronicle

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Environment Testing, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Accreditation/Certification Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Laboratory: Eurofins Lancaster Laboratories Environment Testing, LLC

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
New York	NELAP	10670	04-01-24

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
537 (Mod)	SPE	Water	6:2 Fluorotelomer sulfonic acid
537 (Mod)	SPE	Water	8:2 Fluorotelomer sulfonic acid
537 (Mod)	SPE	Water	Perfluorobutanoic acid
537 (Mod)	SPE	Water	Perfluorodecanesulfonic acid
537 (Mod)	SPE	Water	Perfluoroheptanesulfonic acid
537 (Mod)	SPE	Water	Perfluorooctanesulfonamide
537 (Mod)	SPE	Water	Perfluoropentanoic acid



Method Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Method	Method Description	Protocol	Laboratory
537 (Mod)	EPA 537 Version 1.1 modified	EPA	ELLE
EPA 537.1	EPA 537.1, Ver 1.0 Nov 2018	EPA	ELLE
537.1 DW Prep	Extraction of Perfluorinated Alkyl Acids	EPA	ELLE

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Environment Testing, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300



Sample Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-145941-1
SDG: HOO

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
410-145941-1	GAC INFLUENT	Water	10/05/23 09:30	10/06/23 09:45
410-145941-2	GAC MIDFLUENT	Water	10/05/23 09:35	10/06/23 09:45
410-145941-3	GAC EFFLUENT	Water	10/05/23 09:40	10/06/23 09:45
410-145941-4	FTB01-231005	Water	10/05/23 09:45	10/06/23 09:45
410-145941-5	LTB01-231005	Water	10/05/23 00:00	10/06/23 09:45

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16



Chain of Custody Record

410-145941 Chain of Custody

Client Contact: Jonathan Dippert, <i>Kirk Moline</i>		Sampler: <i>C. Omsby</i>	Lab PM: Tessier, Kelly	Camera Tracking No(s):	COC No: 410-77610-21525.1						
Company: CT Male Associates DPC		Phone:	E-Mail: kelly.tessier@et.eurofinsus.com	State of Origin: <i>NY</i>	Page Page 1 of 2 <i>(2)</i> 1 of 1						
Address: 50 Century Hill Dr		Due Date Requested:	Analysis Requested								
City: Latham		TAT Requested (days): <i>Standards</i>	Preservation Codes:								
State, Zip: NY, 12110		Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No	A - HCL M - Hexane B - NaOH N - None C - Zn Acetate O - AsNaO2 D - Nitric Acid P - Na2O4S E - NaHSO4 Q - Na2SO3 F - MeOH R - Na2S2O3 G - Amchlor S - H2SO4 H - Ascorbic Acid T - TSP Dodecahydrate I - Ice U - Acetone J - DI Water V - MCAA K - EDTA W - pH 4-5 L - EDA Y - Trizma Z - other (specify)								
Project Name: Hoosick Falls WTP		Project #: 41000511	Other:								
Site: <i>14.4756</i>		SSOW#:	Job #:								
Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water, S=solid, O=waste/oil, BT=Tissue, A=Air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	PFC_IDA - (MOD) 7 PFAS Compounds	537_DW - 14 PFAS Drinking Water List	637_DW - 14 PFAS Drinking Water List	Total Number of containers	Special Instructions/Note:
							N	Y	N		
<i>GAC INFLUENT</i>	<i>10/5/23</i>	<i>0930</i>	<i>G</i>	Water	<i>N</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>4</i>	
<i>GAC MIDFLUENT</i>		<i>0935</i>		Water	<i>N</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>4</i>	
<i>GAC EFFLUENT</i>		<i>0940</i>		Water	<i>N</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>4</i>	
<i>FTB01-231005</i>		<i>0945</i>		Water	<i>N</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>4</i>	
<i>LTB01-231005</i>		<i>-</i>		Water	<i>N</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>Y</i>	<i>4</i>	
				Water							
				Water							
				Water							
				Water							
				Water							
				Water							
Possible Hazard Identification <input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological						Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months					
Deliverable Requested: I, II, III, IV, Other (specify) <i>ASP-B, FWH-1 file</i>						Special Instructions/QC Requirements:					
Empty Kit Relinquished by:			Date:		Time:		Method of Shipment:				
Relinquished by: <i>[Signature]</i>			Date/Time: <i>10/5/23 1500</i>		Company: <i>cm</i>		Received by:		Date/Time:		Company:
Relinquished by:			Date/Time:		Company:		Received by:		Date/Time:		Company:
Relinquished by:			Date/Time:		Company:		Received by: <i>[Signature]</i>		Date/Time: <i>10/11/23 0945</i>		Company:
Custody Seals Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No.:				Cooler Temperature(s) °C and Other Remarks: <i>R: 1.5 C: 1.5</i>					



Login Sample Receipt Checklist

Client: CT Male Associates DPC

Job Number: 410-145941-1

SDG Number: HOO

Login Number: 145941

List Source: Eurofins Lancaster Laboratories Environment Testing, LLC

List Number: 1

Creator: Reiff, Nicole L

Question	Answer	Comment
The cooler's custody seal is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature acceptable,where thermal pres is required(</=6C, not frozen).	True	
Cooler Temperature is recorded.	True	
WV:Container Temp acceptable,where thermal pres is required (</=6C, not frozen).	N/A	
WV: Container Temperature is recorded.	N/A	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses.	True	
Is the Field Sampler's name present on COC?	True	
Sample custody seals are intact.	True	
VOA sample vials do not have headspace >6mm in diameter (none, if from WV)?	N/A	

