

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Mr. Kirk Moline  
CT Male Associates DPC  
50 Century Hill Dr  
Latham, New York 12110

Generated 11/17/2023 10:43:14 AM

## JOB DESCRIPTION

Hoosick Falls WTP  
HOO

## JOB NUMBER

410-149823-1


# Eurofins Lancaster Laboratories Environment Testing, LLC

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

Analytical test results meet all requirements of the associated regulatory program (i.e., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis.

## Authorization



Generated  
11/17/2023 10:43:14 AM

---

Authorized for release by  
Kelly Tessier, Project Manager  
[kelly.tessier@et.eurofinsus.com](mailto:kelly.tessier@et.eurofinsus.com)  
(717)205-7820

## Compliance Statement

Analytical test results meet all requirements of the associated regulatory program (e.g., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis. Data qualifiers are applied to note exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- QC results that exceed the upper limits and are associated with non-detect samples are qualified but further narration is not required since the bias is high and does not change a non-detect result. Further narration is also not required with QC blank detection when the associated sample concentration is non-detect or more than ten times the level in the blank.
- Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD is performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.


Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Measurement uncertainty values, as applicable, are available upon request.

Test results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" and tested in the laboratory are not performed within 15 minutes of collection.

This report shall not be reproduced except in full, without the written approval of the laboratory.

**WARRANTY AND LIMITS OF LIABILITY** - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. The foregoing express warranty is exclusive and is given in lieu of all other warranties, expressed or implied, except as otherwise agreed. We disclaim any other warranties, expressed or implied, including a warranty of fitness for particular purpose and warranty of merchantability. In no event shall Eurofins Lancaster Laboratories Environmental, LLC be liable for indirect, special, consequential, or incidental damages including, but not limited to, damages for loss of profit or goodwill regardless of (A) the negligence (either sole or concurrent) of Eurofins Lancaster Laboratories Environmental and (B) whether Eurofins Lancaster Laboratories Environmental has been informed of the possibility of such damages. We accept no legal responsibility for the purposes for which the client uses the test results. Except as otherwise agreed, no purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.



# Table of Contents

Cover Page . . . . .	1	3
Table of Contents . . . . .	4	4
Definitions/Glossary . . . . .	5	5
Case Narrative . . . . .	6	6
Detection Summary . . . . .	7	6
Client Sample Results . . . . .	8	7
Surrogate Summary . . . . .	16	8
Isotope Dilution Summary . . . . .	17	9
QC Sample Results . . . . .	18	9
QC Association Summary . . . . .	21	10
Lab Chronicle . . . . .	23	11
Certification Summary . . . . .	25	12
Method Summary . . . . .	26	12
Sample Summary . . . . .	27	13
Chain of Custody . . . . .	28	14
Receipt Checklists . . . . .	29	15
		16

## Definitions/Glossary

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

### Qualifiers

#### LCMS

Qualifier	Qualifier Description
*5-	Isotope dilution analyte is outside acceptance limits, low biased.
cn	Refer to Case Narrative for further detail
U	Indicates the analyte was analyzed for but not detected.

### Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
□	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

## Case Narrative

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

### Job ID: 410-149823-1

Laboratory: Eurofins Lancaster Laboratories Environment Testing, LLC

#### Narrative

##### Job Narrative 410-149823-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers are applied to indicate exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method. Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

#### Receipt

The samples were received on 11/3/2023 9:45 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 0.9°C

#### PFAS

Method 537.1\_DW: The following sample was found to contain residual chlorine: GAC Influent (410-149823-1).

Method PFC\_IDA: The recovery for labeled isotope: M2-8:2 FTS is outside the QC acceptance limits in the opening continuing calibration verification standard. Since the recovery for the labeled isotope is within QC limits in the following samples: GAC Influent (410-149823-1) and GAC Midfluent (410-149823-2), the data is reported.

Method PFC\_IDA: The recovery for 13C4 PFBA in GAC Influent (410-149823-1) is below the QC acceptance limit. The client was contacted and the data reported.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

## Detection Summary

Client: CT Male Associates DPC  
 Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
 SDG: HOO

### **Client Sample ID: GAC Influent**

**Lab Sample ID: 410-149823-1**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	4.3		1.9	ng/L	1	537 (Mod)	Total/NA	
Perfluoroctanesulfonamide	2.7		1.9	ng/L	1	537 (Mod)	Total/NA	
Perfluoropentanoic acid	2.6		1.9	ng/L	1	537 (Mod)	Total/NA	
Perfluoroheptanoic acid	9.3	cn	1.8	ng/L	1	EPA 537.1	Total/NA	
Perfluorohexanoic acid	8.9	cn	1.8	ng/L	1	EPA 537.1	Total/NA	
Perfluoroctanesulfonic acid	3.4	cn	1.8	ng/L	1	EPA 537.1	Total/NA	
Perfluoroctanoic acid - DL	410	cn	18	ng/L	10	EPA 537.1	Total/NA	

### **Client Sample ID: GAC Midfluent**

**Lab Sample ID: 410-149823-2**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	4.4		1.8	ng/L	1	537 (Mod)	Total/NA	

### **Client Sample ID: GAC Effluent**

**Lab Sample ID: 410-149823-3**

No Detections.

### **Client Sample ID: PV-2\_25**

**Lab Sample ID: 410-149823-4**

No Detections.

### **Client Sample ID: PV-2\_50**

**Lab Sample ID: 410-149823-5**

No Detections.

### **Client Sample ID: PV-2\_75**

**Lab Sample ID: 410-149823-6**

No Detections.

### **Client Sample ID: FTB01-231102**

**Lab Sample ID: 410-149823-7**

No Detections.

### **Client Sample ID: LTB01-231102**

**Lab Sample ID: 410-149823-8**

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Environment Testing, LLC

# Client Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

## Client Sample ID: GAC Influent

Date Collected: 11/02/23 09:50  
Date Received: 11/03/23 09:45

## Lab Sample ID: 410-149823-1

Matrix: Water

### Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.9	U	1.9	ng/L		11/07/23 06:36	11/15/23 04:57	1
8:2 Fluorotelomer sulfonic acid	1.9	U	1.9	ng/L		11/07/23 06:36	11/15/23 04:57	1
<b>Perfluorobutanoic acid</b>	<b>4.3</b>		1.9	ng/L		11/07/23 06:36	11/15/23 04:57	1
Perfluorodecanesulfonic acid	1.9	U	1.9	ng/L		11/07/23 06:36	11/15/23 04:57	1
Perfluoroheptanesulfonic acid	1.9	U	1.9	ng/L		11/07/23 06:36	11/15/23 04:57	1
<b>Perfluorooctanesulfonamide</b>	<b>2.7</b>		1.9	ng/L		11/07/23 06:36	11/15/23 04:57	1
<b>Perfluoropentanoic acid</b>	<b>2.6</b>		1.9	ng/L		11/07/23 06:36	11/15/23 04:57	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	110		40 - 200			11/07/23 06:36	11/15/23 04:57	1
M2-8:2 FTS	106	cn	37 - 200			11/07/23 06:36	11/15/23 04:57	1
13C4 PFBA	14	*5- cn	22 - 174			11/07/23 06:36	11/15/23 04:57	1
13C5 PFPeA	82		33 - 196			11/07/23 06:36	11/15/23 04:57	1
13C8 PFOS	95		59 - 155			11/07/23 06:36	11/15/23 04:57	1
13C8 FOSA	73		10 - 155			11/07/23 06:36	11/15/23 04:57	1
13C3 PFHxS	116		48 - 169			11/07/23 06:36	11/15/23 04:57	1

### Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
NMeFOSAA	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Perfluorobutanesulfonic acid	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Perfluorodecanoic acid	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Perfluorododecanoic acid	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
<b>Perfluoroheptanoic acid</b>	<b>9.3</b> cn		1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Perfluorohexanesulfonic acid	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
<b>Perfluorohexanoic acid</b>	<b>8.9</b> cn		1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Perfluorononanoic acid	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
<b>Perfluorooctanesulfonic acid</b>	<b>3.4</b> cn		1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Perfluorotetradecanoic acid	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Perfluorotridecanoic acid	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Perfluoroundecanoic acid	1.8	U cn	1.8	ng/L		11/07/23 09:19	11/08/23 13:06	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	104	cn	70 - 130			11/07/23 09:19	11/08/23 13:06	1
13C2 PFHxA	110	cn	70 - 130			11/07/23 09:19	11/08/23 13:06	1
d5-NEtFOSAA	109	cn	70 - 130			11/07/23 09:19	11/08/23 13:06	1

### Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018 - DL

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
<b>Perfluorooctanoic acid</b>	<b>410</b> cn		18	ng/L		11/07/23 09:19	11/09/23 21:44	10
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	117	cn	70 - 130			11/07/23 09:19	11/09/23 21:44	10
13C2 PFHxA	122	cn	70 - 130			11/07/23 09:19	11/09/23 21:44	10
d5-NEtFOSAA	114	cn	70 - 130			11/07/23 09:19	11/09/23 21:44	10

# Client Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

## Client Sample ID: GAC Midfluent

Date Collected: 11/02/23 09:55  
Date Received: 11/03/23 09:45

## Lab Sample ID: 410-149823-2

Matrix: Water

### Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:08	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:08	1
<b>Perfluorobutanoic acid</b>	<b>4.4</b>		1.8	ng/L		11/07/23 06:36	11/15/23 05:08	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:08	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:08	1
Perfluoroctanesulfonamide	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:08	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:08	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	95		40 - 200			11/07/23 06:36	11/15/23 05:08	1
M2-8:2 FTS	102	cn	37 - 200			11/07/23 06:36	11/15/23 05:08	1
13C4 PFBA	29		22 - 174			11/07/23 06:36	11/15/23 05:08	1
13C5 PFPeA	112		33 - 196			11/07/23 06:36	11/15/23 05:08	1
13C8 PFOS	97		59 - 155			11/07/23 06:36	11/15/23 05:08	1
13C8 FOSA	82		10 - 155			11/07/23 06:36	11/15/23 05:08	1
13C3 PFHxS	104		48 - 169			11/07/23 06:36	11/15/23 05:08	1

### Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
NMeFOSAA	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluorobutanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluorodecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluorododecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluoroheptanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluorohexanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluorohexanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluorononanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluoroctanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluoroctanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluorotetradecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluorotridecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Perfluoroundecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:18	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	110		70 - 130			11/07/23 09:19	11/08/23 13:18	1
13C2 PFHxA	115		70 - 130			11/07/23 09:19	11/08/23 13:18	1
d5-NEtFOSAA	108		70 - 130			11/07/23 09:19	11/08/23 13:18	1

# Client Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

## Client Sample ID: GAC Effluent

Date Collected: 11/02/23 10:00  
Date Received: 11/03/23 09:45

## Lab Sample ID: 410-149823-3

Matrix: Water

### Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:30	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:30	1
Perfluorobutanoic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:30	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:30	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:30	1
Perfluoroctanesulfonamide	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:30	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 05:30	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	100		40 - 200			11/07/23 06:36	11/15/23 05:30	1
M2-8:2 FTS	106		37 - 200			11/07/23 06:36	11/15/23 05:30	1
13C4 PFBA	43		22 - 174			11/07/23 06:36	11/15/23 05:30	1
13C5 PFPeA	112		33 - 196			11/07/23 06:36	11/15/23 05:30	1
13C8 PFOS	106		59 - 155			11/07/23 06:36	11/15/23 05:30	1
13C8 FOSA	86		10 - 155			11/07/23 06:36	11/15/23 05:30	1
13C3 PFHxS	107		48 - 169			11/07/23 06:36	11/15/23 05:30	1

### Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
NMeFOSAA	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluorohexanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluooctanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluooctanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 13:29	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	97		70 - 130			11/07/23 09:19	11/08/23 13:29	1
13C2 PFHxA	108		70 - 130			11/07/23 09:19	11/08/23 13:29	1
d5-NEtFOSAA	106		70 - 130			11/07/23 09:19	11/08/23 13:29	1

# Client Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1

SDG: HOO

**Client Sample ID: PV-2\_25**

Date Collected: 11/02/23 10:05

Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-4**

Matrix: Water

**Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/15/23 05:41	1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/15/23 05:41	1
Perfluorobutanoic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/15/23 05:41	1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/15/23 05:41	1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/15/23 05:41	1
Perfluoroctanesulfonamide	1.7	U	1.7	ng/L		11/07/23 06:36	11/15/23 05:41	1
Perfluoropentanoic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/15/23 05:41	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	98		40 - 200			11/07/23 06:36	11/15/23 05:41	1
M2-8:2 FTS	109		37 - 200			11/07/23 06:36	11/15/23 05:41	1
13C4 PFBA	28		22 - 174			11/07/23 06:36	11/15/23 05:41	1
13C5 PFPeA	96		33 - 196			11/07/23 06:36	11/15/23 05:41	1
13C8 PFOS	99		59 - 155			11/07/23 06:36	11/15/23 05:41	1
13C8 FOSA	82		10 - 155			11/07/23 06:36	11/15/23 05:41	1
13C3 PFHxS	109		48 - 169			11/07/23 06:36	11/15/23 05:41	1

**Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
NMeFOSAA	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluorobutanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluorodecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluorododecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluoroheptanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluorohexanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluorohexanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluorononanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluooctanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluooctanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluorotetradecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluorotridecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Perfluoroundecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:41	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	110		70 - 130			11/07/23 09:19	11/08/23 13:41	1
13C2 PFHxA	107		70 - 130			11/07/23 09:19	11/08/23 13:41	1
d5-NEtFOSAA	97		70 - 130			11/07/23 09:19	11/08/23 13:41	1

# Client Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1

SDG: HOO

**Client Sample ID: PV-2\_50**

Date Collected: 11/02/23 10:07

Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-5**

Matrix: Water

## Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/16/23 13:14	1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/16/23 13:14	1
Perfluorobutanoic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/16/23 13:14	1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/16/23 13:14	1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/16/23 13:14	1
Perfluoroctanesulfonamide	1.7	U	1.7	ng/L		11/07/23 06:36	11/16/23 13:14	1
Perfluoropentanoic acid	1.7	U	1.7	ng/L		11/07/23 06:36	11/16/23 13:14	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	111		40 - 200			11/07/23 06:36	11/16/23 13:14	1
M2-8:2 FTS	104		37 - 200			11/07/23 06:36	11/16/23 13:14	1
13C4 PFBA	37		22 - 174			11/07/23 06:36	11/16/23 13:14	1
13C5 PFPeA	140		33 - 196			11/07/23 06:36	11/16/23 13:14	1
13C8 PFOS	137		59 - 155			11/07/23 06:36	11/16/23 13:14	1
13C8 FOSA	112		10 - 155			11/07/23 06:36	11/16/23 13:14	1
13C3 PFHxS	153		48 - 169			11/07/23 06:36	11/16/23 13:14	1

## Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
NMeFOSAA	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluorobutanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluorodecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluorododecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluoroheptanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluorohexanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluorohexanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluorononanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluooctanesulfonic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluooctanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluorotetradecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluorotridecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Perfluoroundecanoic acid	1.7	U	1.7	ng/L		11/07/23 09:19	11/08/23 13:52	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	114		70 - 130			11/07/23 09:19	11/08/23 13:52	1
13C2 PFHxA	108		70 - 130			11/07/23 09:19	11/08/23 13:52	1
d5-NEtFOSAA	97		70 - 130			11/07/23 09:19	11/08/23 13:52	1

# Client Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1

SDG: HOO

**Client Sample ID: PV-2\_75**

Date Collected: 11/02/23 10:10

Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-6**

Matrix: Water

**Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L	11/07/23 06:36	11/15/23 06:03		1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L	11/07/23 06:36	11/15/23 06:03		1
Perfluorobutanoic acid	1.8	U	1.8	ng/L	11/07/23 06:36	11/15/23 06:03		1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L	11/07/23 06:36	11/15/23 06:03		1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L	11/07/23 06:36	11/15/23 06:03		1
Perfluoroctanesulfonamide	1.8	U	1.8	ng/L	11/07/23 06:36	11/15/23 06:03		1
Perfluoropentanoic acid	1.8	U	1.8	ng/L	11/07/23 06:36	11/15/23 06:03		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	101		40 - 200			11/07/23 06:36	11/15/23 06:03	1
M2-8:2 FTS	102		37 - 200			11/07/23 06:36	11/15/23 06:03	1
13C4 PFBA	39		22 - 174			11/07/23 06:36	11/15/23 06:03	1
13C5 PFPeA	97		33 - 196			11/07/23 06:36	11/15/23 06:03	1
13C8 PFOS	98		59 - 155			11/07/23 06:36	11/15/23 06:03	1
13C8 FOSA	84		10 - 155			11/07/23 06:36	11/15/23 06:03	1
13C3 PFHxS	110		48 - 169			11/07/23 06:36	11/15/23 06:03	1

**Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
NMeFOSAA	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluorodecanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluorododecanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluorohexanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluorononanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluooctanesulfonic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluooctanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L	11/07/23 09:19	11/08/23 14:04		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	111		70 - 130			11/07/23 09:19	11/08/23 14:04	1
13C2 PFHxA	111		70 - 130			11/07/23 09:19	11/08/23 14:04	1
d5-NEtFOSAA	104		70 - 130			11/07/23 09:19	11/08/23 14:04	1

# Client Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

**Client Sample ID: FTB01-231102**

**Lab Sample ID: 410-149823-7**

**Matrix: Water**

Date Collected: 11/02/23 10:15  
Date Received: 11/03/23 09:45

## Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/16/23 13:25	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/16/23 13:25	1
Perfluorobutanoic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/16/23 13:25	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/16/23 13:25	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/16/23 13:25	1
Perfluoroctanesulfonamide	1.8	U	1.8	ng/L		11/07/23 06:36	11/16/23 13:25	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/16/23 13:25	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	106		40 - 200			11/07/23 06:36	11/16/23 13:25	1
M2-8:2 FTS	107		37 - 200			11/07/23 06:36	11/16/23 13:25	1
13C4 PFBA	71		22 - 174			11/07/23 06:36	11/16/23 13:25	1
13C5 PFPeA	112		33 - 196			11/07/23 06:36	11/16/23 13:25	1
13C8 PFOS	138		59 - 155			11/07/23 06:36	11/16/23 13:25	1
13C8 FOSA	123		10 - 155			11/07/23 06:36	11/16/23 13:25	1
13C3 PFHxS	155		48 - 169			11/07/23 06:36	11/16/23 13:25	1

## Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
NMeFOSAA	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluorohexanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluooctanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluooctanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:16	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	108		70 - 130			11/07/23 09:19	11/08/23 14:16	1
13C2 PFHxA	113		70 - 130			11/07/23 09:19	11/08/23 14:16	1
d5-NEtFOSAA	101		70 - 130			11/07/23 09:19	11/08/23 14:16	1

# Client Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1

SDG: HOO

**Client Sample ID: LTB01-231102**

**Lab Sample ID: 410-149823-8**

**Matrix: Water**

Date Collected: 11/02/23 00:00

Date Received: 11/03/23 09:45

## Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 06:25	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 06:25	1
Perfluorobutanoic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 06:25	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 06:25	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 06:25	1
Perfluoroctanesulfonamide	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 06:25	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		11/07/23 06:36	11/15/23 06:25	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	97		40 - 200			11/07/23 06:36	11/15/23 06:25	1
M2-8:2 FTS	113		37 - 200			11/07/23 06:36	11/15/23 06:25	1
13C4 PFBA	67		22 - 174			11/07/23 06:36	11/15/23 06:25	1
13C5 PFPeA	107		33 - 196			11/07/23 06:36	11/15/23 06:25	1
13C8 PFOS	101		59 - 155			11/07/23 06:36	11/15/23 06:25	1
13C8 FOSA	83		10 - 155			11/07/23 06:36	11/15/23 06:25	1
13C3 PFHxS	111		48 - 169			11/07/23 06:36	11/15/23 06:25	1

## Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
NMeFOSAA	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluorohexanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluooctanesulfonic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluooctanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		11/07/23 09:19	11/08/23 14:27	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C2 PFDA	110		70 - 130			11/07/23 09:19	11/08/23 14:27	1
13C2 PFHxA	104		70 - 130			11/07/23 09:19	11/08/23 14:27	1
d5-NEtFOSAA	104		70 - 130			11/07/23 09:19	11/08/23 14:27	1

## Surrogate Summary

Client: CT Male Associates DPC  
 Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
 SDG: HOO

**Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018**

**Matrix: Water**

**Prep Type: Total/NA**

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		PFDA (70-130)	PFHxA (70-130)	d5NEFOS (70-130)
410-149823-1	GAC Influent	104 cn	110 cn	109 cn
410-149823-1 - DL	GAC Influent	117 cn	122 cn	114 cn
410-149823-2	GAC Midfluent	110	115	108
410-149823-3	GAC Effluent	97	108	106
410-149823-4	PV-2_25	110	107	97
410-149823-5	PV-2_50	114	108	97
410-149823-6	PV-2_75	111	111	104
410-149823-7	FTB01-231102	108	113	101
410-149823-8	LTB01-231102	110	104	104
LCS 410-440446/2-A	Lab Control Sample	103	111	112
LCSD 410-440446/3-A	Lab Control Sample Dup	100	112	103
MB 410-440446/1-A	Method Blank	104	108	107

**Surrogate Legend**

PFDA = 13C2 PFDA

PFHxA = 13C2 PFHxA

d5NEFOS = d5-NEtFOSAA

# Isotope Dilution Summary

Client: CT Male Associates DPC  
 Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
 SDG: HOO

## Method: 537 (Mod) - EPA 537 Version 1.1 modified

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Isotope Dilution Recovery (Acceptance Limits)						
		M262FTS (40-200)	M282FTS (37-200)	PFBA (22-174)	PFPeA (33-196)	C8PFOS (59-155)	PFOSA (10-155)	C3PFHS (48-169)
410-149823-1	GAC Influent	110	106 cn	14 *5- cn	82	95	73	116
410-149823-2	GAC Midfluent	95	102 cn	29	112	97	82	104
410-149823-3	GAC Effluent	100	106	43	112	106	86	107
410-149823-4	PV-2_25	98	109	28	96	99	82	109
410-149823-5	PV-2_50	111	104	37	140	137	112	153
410-149823-6	PV-2_75	101	102	39	97	98	84	110
410-149823-7	FTB01-231102	106	107	71	112	138	123	155
410-149823-8	LTB01-231102	97	113	67	107	101	83	111
LCS 410-440311/2-A	Lab Control Sample	110	96	47	115	126	102	124
LCSD 410-440311/3-A	Lab Control Sample Dup	120	119	43	101	101	89	107
MB 410-440311/1-A	Method Blank	112	102	99	133	135	108	152

### Surrogate Legend

M262FTS = M2-6:2 FTS

M282FTS = M2-8:2 FTS

PFBA = 13C4 PFBA

PFPeA = 13C5 PFPeA

C8PFOS = 13C8 PFOS

PFOSA = 13C8 FOSA

C3PFHS = 13C3 PFHxS

# QC Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

## Method: 537 (Mod) - EPA 537 Version 1.1 modified

**Lab Sample ID:** MB 410-440311/1-A

**Client Sample ID:** Method Blank

**Matrix:** Water

**Prep Type:** Total/NA

**Analysis Batch:** 444192

**Prep Batch:** 440311

Analyte	MB	MB	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier								
6:2 Fluorotelomer sulfonic acid	2.0	U	2.0		2.0	ng/L		11/07/23 06:36	11/16/23 12:19	1
8:2 Fluorotelomer sulfonic acid	2.0	U	2.0		2.0	ng/L		11/07/23 06:36	11/16/23 12:19	1
Perfluorobutanoic acid	2.0	U	2.0		2.0	ng/L		11/07/23 06:36	11/16/23 12:19	1
Perfluorodecanesulfonic acid	2.0	U	2.0		2.0	ng/L		11/07/23 06:36	11/16/23 12:19	1
Perfluoroheptanesulfonic acid	2.0	U	2.0		2.0	ng/L		11/07/23 06:36	11/16/23 12:19	1
Perfluorooctanesulfonamide	2.0	U	2.0		2.0	ng/L		11/07/23 06:36	11/16/23 12:19	1
Perfluoropentanoic acid	2.0	U	2.0		2.0	ng/L		11/07/23 06:36	11/16/23 12:19	1
<hr/>										
Isotope Dilution	MB	MB	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac	10	
	Result	Qualifier								
M2-6:2 FTS	112		40 - 200			11/07/23 06:36	11/16/23 12:19	1	10	
M2-8:2 FTS	102		37 - 200			11/07/23 06:36	11/16/23 12:19	1	11	
13C4 PFBA	99		22 - 174			11/07/23 06:36	11/16/23 12:19	1	12	
13C5 PFPeA	133		33 - 196			11/07/23 06:36	11/16/23 12:19	1	13	
13C8 PFOS	135		59 - 155			11/07/23 06:36	11/16/23 12:19	1	14	
13C8 FOSA	108		10 - 155			11/07/23 06:36	11/16/23 12:19	1	15	
13C3 PFHxS	152		48 - 169			11/07/23 06:36	11/16/23 12:19	1	16	

**Lab Sample ID:** LCS 410-440311/2-A

**Client Sample ID:** Lab Control Sample

**Matrix:** Water

**Prep Type:** Total/NA

**Analysis Batch:** 444192

**Prep Batch:** 440311

Analyte	Spike	LCS	LCS	Result	Qualifier	Unit	D	%Rec	Limits	%Rec
	Added	Result	Qualifier							
6:2 Fluorotelomer sulfonic acid		24.3		17.9		ng/L		74	61 - 132	
8:2 Fluorotelomer sulfonic acid		24.5		20.8		ng/L		85	55 - 134	
Perfluorobutanoic acid		25.6		17.8		ng/L		69	58 - 130	
Perfluorodecanesulfonic acid		24.7		19.0		ng/L		77	55 - 130	
Perfluoroheptanesulfonic acid		24.4		19.4		ng/L		80	59 - 130	
Perfluorooctanesulfonamide		25.6		21.1		ng/L		82	67 - 132	
Perfluoropentanoic acid		25.6		18.5		ng/L		72	60 - 130	
<hr/>										
Isotope Dilution	LCS	LCS	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac	14	
	Result	Qualifier								
M2-6:2 FTS	110		40 - 200			11/07/23 06:36	11/16/23 12:19	1	15	
M2-8:2 FTS	96		37 - 200			11/07/23 06:36	11/16/23 12:19	1	16	
13C4 PFBA	47		22 - 174			11/07/23 06:36	11/16/23 12:19	1		
13C5 PFPeA	115		33 - 196			11/07/23 06:36	11/16/23 12:19	1		
13C8 PFOS	126		59 - 155			11/07/23 06:36	11/16/23 12:19	1		
13C8 FOSA	102		10 - 155			11/07/23 06:36	11/16/23 12:19	1		
13C3 PFHxS	124		48 - 169			11/07/23 06:36	11/16/23 12:19	1		

**Lab Sample ID:** LCSD 410-440311/3-A

**Client Sample ID:** Lab Control Sample Dup

**Matrix:** Water

**Prep Type:** Total/NA

**Analysis Batch:** 443245

**Prep Batch:** 440311

Analyte	Spike	LCSD	LCSD	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit
	Added	Result	Qualifier								
6:2 Fluorotelomer sulfonic acid		24.3		17.5		ng/L		72	61 - 132	0	30
8:2 Fluorotelomer sulfonic acid		24.5		16.9		ng/L		69	55 - 134	14	30
Perfluorobutanoic acid		25.6		17.5		ng/L		68	58 - 130	1	30
Perfluorodecanesulfonic acid		24.7		20.9		ng/L		85	55 - 130	1	30
Perfluoroheptanesulfonic acid		24.4		18.5		ng/L		76	59 - 130	1	30

Eurofins Lancaster Laboratories Environment Testing, LLC

# QC Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

## Method: 537 (Mod) - EPA 537 Version 1.1 modified (Continued)

**Lab Sample ID:** LCSD 410-440311/3-A

**Client Sample ID:** Lab Control Sample Dup

**Matrix:** Water

**Prep Type:** Total/NA

**Analysis Batch:** 443245

**Prep Batch:** 440311

Analyte		Spike	LCSD	LCSD	Unit	D	%Rec	%Rec	RPD	RPD	Limit
		Added	Result	Qualifier							
Perfluoroctanesulfonamide		25.6	21.8		ng/L		85	67 - 132	1		30
Perfluoropentanoic acid		25.6	17.9		ng/L		70	60 - 130	4		30
<b>Isotope Dilution</b>											
M2-6:2 FTS	%Recovery	LCSD	LCSD	Limits							
M2-8:2 FTS											
13C4 PFBA											
13C5 PFPeA											
13C8 PFOS											
13C8 FOSA											
13C3 PFHxS											

## Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

**Lab Sample ID:** MB 410-440446/1-A

**Client Sample ID:** Method Blank

**Matrix:** Water

**Prep Type:** Total/NA

**Analysis Batch:** 441016

**Prep Batch:** 440446

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac			
	Result	Qualifier									
NEtFOSAA	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
NMeFOSAA	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluorobutanesulfonic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluorodecanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluorododecanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluoroheptanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluorohexanesulfonic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluorohexanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluorononanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluooctanesulfonic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluooctanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluorotetradecanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluorotridecanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
Perfluoroundecanoic acid	2.0	U	2.0	ng/L		11/07/23 09:19	11/08/23 12:20	1			
<b>Surrogate</b>											
	MB	MB									
	%Recovery	Qualifier	Limits								
13C2 PFDA	104		70 - 130								
13C2 PFHxA	108		70 - 130								
d5-NEtFOSAA	107		70 - 130								

**Lab Sample ID:** LCS 410-440446/2-A

**Client Sample ID:** Lab Control Sample

**Matrix:** Water

**Prep Type:** Total/NA

**Analysis Batch:** 441016

**Prep Batch:** 440446

Analyte	Spike	LCS	LCS	Unit	D	%Rec	%Rec	RPD	RPD	Limit
	Added	Result	Qualifier							
NEtFOSAA	20.5	18.2		ng/L		89	70 - 130			
NMeFOSAA	20.5	19.7		ng/L		96	70 - 130			
Perfluorobutanesulfonic acid	18.1	21.6		ng/L		119	70 - 130			
Perfluorodecanoic acid	20.5	21.1		ng/L		103	70 - 130			
Perfluorododecanoic acid	20.5	20.5		ng/L		100	70 - 130			

# QC Sample Results

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

## Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018 (Continued)

**Lab Sample ID:** LCS 410-440446/2-A

**Matrix:** Water

**Analysis Batch:** 441016

**Client Sample ID:** Lab Control Sample

**Prep Type:** Total/NA

**Prep Batch:** 440446

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
Perfluoroheptanoic acid	20.5	22.4		ng/L	109	70 - 130	
Perfluorohexanesulfonic acid	18.7	22.8		ng/L	122	70 - 130	
Perfluorohexanoic acid	20.5	22.8		ng/L	111	70 - 130	
Perfluorononanoic acid	20.5	20.1		ng/L	98	70 - 130	
Perfluoroctanesulfonic acid	19.0	19.7		ng/L	104	70 - 130	
Perfluoroctanoic acid	20.5	21.1		ng/L	103	70 - 130	
Perfluorotetradecanoic acid	20.5	22.5		ng/L	110	70 - 130	
Perfluorotridecanoic acid	20.5	19.5		ng/L	95	70 - 130	
Perfluoroundecanoic acid	20.5	20.6		ng/L	101	70 - 130	

Surrogate	LCS %Recovery	LCS Qualifier	Limits
13C2 PFDA	103		70 - 130
13C2 PFHxA	111		70 - 130
d5-NEtFOSAA	112		70 - 130

**Lab Sample ID:** LCSD 410-440446/3-A

**Matrix:** Water

**Analysis Batch:** 441016

**Client Sample ID:** Lab Control Sample Dup

**Prep Type:** Total/NA

**Prep Batch:** 440446

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	RPD	Limit
NEtFOSAA	20.5	20.1		ng/L	98	70 - 130	10	30
NMeFOSAA	20.5	19.7		ng/L	96	70 - 130	0	30
Perfluorobutanesulfonic acid	18.1	19.1		ng/L	105	70 - 130	12	30
Perfluorodecanoic acid	20.5	20.9		ng/L	102	70 - 130	1	30
Perfluorododecanoic acid	20.5	22.2		ng/L	109	70 - 130	8	30
Perfluoroheptanoic acid	20.5	21.1		ng/L	103	70 - 130	6	30
Perfluorohexanesulfonic acid	18.7	18.6		ng/L	100	70 - 130	20	30
Perfluorohexanoic acid	20.5	22.9		ng/L	112	70 - 130	1	30
Perfluorononanoic acid	20.5	20.4		ng/L	99	70 - 130	1	30
Perfluoroctanesulfonic acid	19.0	19.1		ng/L	101	70 - 130	3	30
Perfluoroctanoic acid	20.5	22.6		ng/L	110	70 - 130	7	30
Perfluorotetradecanoic acid	20.5	23.1		ng/L	113	70 - 130	3	30
Perfluorotridecanoic acid	20.5	20.1		ng/L	98	70 - 130	3	30
Perfluoroundecanoic acid	20.5	19.6		ng/L	96	70 - 130	5	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
13C2 PFDA	100		70 - 130
13C2 PFHxA	112		70 - 130
d5-NEtFOSAA	103		70 - 130

# QC Association Summary

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

## LCMS

### Prep Batch: 440311

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-149823-1	GAC Influent	Total/NA	Water	SPE	1
410-149823-2	GAC Midfluent	Total/NA	Water	SPE	2
410-149823-3	GAC Effluent	Total/NA	Water	SPE	3
410-149823-4	PV-2_25	Total/NA	Water	SPE	4
410-149823-5	PV-2_50	Total/NA	Water	SPE	5
410-149823-6	PV-2_75	Total/NA	Water	SPE	6
410-149823-7	FTB01-231102	Total/NA	Water	SPE	7
410-149823-8	LTB01-231102	Total/NA	Water	SPE	8
MB 410-440311/1-A	Method Blank	Total/NA	Water	SPE	9
LCS 410-440311/2-A	Lab Control Sample	Total/NA	Water	SPE	10
LCSD 410-440311/3-A	Lab Control Sample Dup	Total/NA	Water	SPE	

### Prep Batch: 440446

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-149823-1	GAC Influent	Total/NA	Water	537.1 DW Prep	11
410-149823-1 - DL	GAC Influent	Total/NA	Water	537.1 DW Prep	12
410-149823-2	GAC Midfluent	Total/NA	Water	537.1 DW Prep	13
410-149823-3	GAC Effluent	Total/NA	Water	537.1 DW Prep	14
410-149823-4	PV-2_25	Total/NA	Water	537.1 DW Prep	15
410-149823-5	PV-2_50	Total/NA	Water	537.1 DW Prep	16
410-149823-6	PV-2_75	Total/NA	Water	537.1 DW Prep	
410-149823-7	FTB01-231102	Total/NA	Water	537.1 DW Prep	
410-149823-8	LTB01-231102	Total/NA	Water	537.1 DW Prep	
MB 410-440446/1-A	Method Blank	Total/NA	Water	537.1 DW Prep	
LCS 410-440446/2-A	Lab Control Sample	Total/NA	Water	537.1 DW Prep	
LCSD 410-440446/3-A	Lab Control Sample Dup	Total/NA	Water	537.1 DW Prep	

### Analysis Batch: 441016

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-149823-1	GAC Influent	Total/NA	Water	EPA 537.1	440446
410-149823-2	GAC Midfluent	Total/NA	Water	EPA 537.1	440446
410-149823-3	GAC Effluent	Total/NA	Water	EPA 537.1	440446
410-149823-4	PV-2_25	Total/NA	Water	EPA 537.1	440446
410-149823-5	PV-2_50	Total/NA	Water	EPA 537.1	440446
410-149823-6	PV-2_75	Total/NA	Water	EPA 537.1	440446
410-149823-7	FTB01-231102	Total/NA	Water	EPA 537.1	440446
410-149823-8	LTB01-231102	Total/NA	Water	EPA 537.1	440446
MB 410-440446/1-A	Method Blank	Total/NA	Water	EPA 537.1	440446
LCS 410-440446/2-A	Lab Control Sample	Total/NA	Water	EPA 537.1	440446
LCSD 410-440446/3-A	Lab Control Sample Dup	Total/NA	Water	EPA 537.1	440446

### Analysis Batch: 441655

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-149823-1 - DL	GAC Influent	Total/NA	Water	EPA 537.1	440446

### Analysis Batch: 443245

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-149823-1	GAC Influent	Total/NA	Water	537 (Mod)	440311
410-149823-2	GAC Midfluent	Total/NA	Water	537 (Mod)	440311
410-149823-3	GAC Effluent	Total/NA	Water	537 (Mod)	440311
410-149823-4	PV-2_25	Total/NA	Water	537 (Mod)	440311

# QC Association Summary

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

## LCMS (Continued)

### Analysis Batch: 443245 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-149823-6	PV-2_75	Total/NA	Water	537 (Mod)	440311
410-149823-8	LTB01-231102	Total/NA	Water	537 (Mod)	440311
LCSD 410-440311/3-A	Lab Control Sample Dup	Total/NA	Water	537 (Mod)	440311

### Analysis Batch: 444192

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-149823-5	PV-2_50	Total/NA	Water	537 (Mod)	440311
410-149823-7	FTB01-231102	Total/NA	Water	537 (Mod)	440311
MB 410-440311/1-A	Method Blank	Total/NA	Water	537 (Mod)	440311
LCS 410-440311/2-A	Lab Control Sample	Total/NA	Water	537 (Mod)	440311

## Lab Chronicle

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

### Client Sample ID: GAC Influent

Date Collected: 11/02/23 09:50  
Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-1**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			440311	RC3V	ELLE	11/07/23 06:36
Total/NA	Analysis	537 (Mod)		1	443245	DS2G	ELLE	11/15/23 04:57
Total/NA	Prep	537.1 DW Prep			440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1		1	441016	WR4P	ELLE	11/08/23 13:06
Total/NA	Prep	537.1 DW Prep	DL		440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1	DL	10	441655	WR4P	ELLE	11/09/23 21:44

### Client Sample ID: GAC Midfluent

Date Collected: 11/02/23 09:55  
Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-2**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			440311	RC3V	ELLE	11/07/23 06:36
Total/NA	Analysis	537 (Mod)		1	443245	DS2G	ELLE	11/15/23 05:08
Total/NA	Prep	537.1 DW Prep			440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1		1	441016	WR4P	ELLE	11/08/23 13:18

### Client Sample ID: GAC Effluent

Date Collected: 11/02/23 10:00  
Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-3**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			440311	RC3V	ELLE	11/07/23 06:36
Total/NA	Analysis	537 (Mod)		1	443245	DS2G	ELLE	11/15/23 05:30
Total/NA	Prep	537.1 DW Prep			440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1		1	441016	WR4P	ELLE	11/08/23 13:29

### Client Sample ID: PV-2\_25

Date Collected: 11/02/23 10:05  
Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-4**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			440311	RC3V	ELLE	11/07/23 06:36
Total/NA	Analysis	537 (Mod)		1	443245	DS2G	ELLE	11/15/23 05:41
Total/NA	Prep	537.1 DW Prep			440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1		1	441016	WR4P	ELLE	11/08/23 13:41

### Client Sample ID: PV-2\_50

Date Collected: 11/02/23 10:07  
Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-5**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			440311	RC3V	ELLE	11/07/23 06:36
Total/NA	Analysis	537 (Mod)		1	444192	QD9Y	ELLE	11/16/23 13:14
Total/NA	Prep	537.1 DW Prep			440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1		1	441016	WR4P	ELLE	11/08/23 13:52

## Lab Chronicle

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1

SDG: HOO

### **Client Sample ID: PV-2\_75**

Date Collected: 11/02/23 10:10

Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-6**

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			440311	RC3V	ELLE	11/07/23 06:36
Total/NA	Analysis	537 (Mod)		1	443245	DS2G	ELLE	11/15/23 06:03
Total/NA	Prep	537.1 DW Prep			440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1		1	441016	WR4P	ELLE	11/08/23 14:04

### **Client Sample ID: FTB01-231102**

Date Collected: 11/02/23 10:15

Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-7**

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			440311	RC3V	ELLE	11/07/23 06:36
Total/NA	Analysis	537 (Mod)		1	444192	QD9Y	ELLE	11/16/23 13:25
Total/NA	Prep	537.1 DW Prep			440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1		1	441016	WR4P	ELLE	11/08/23 14:16

### **Client Sample ID: LTB01-231102**

Date Collected: 11/02/23 00:00

Date Received: 11/03/23 09:45

**Lab Sample ID: 410-149823-8**

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			440311	RC3V	ELLE	11/07/23 06:36
Total/NA	Analysis	537 (Mod)		1	443245	DS2G	ELLE	11/15/23 06:25
Total/NA	Prep	537.1 DW Prep			440446	HQ8B	ELLE	11/07/23 09:19
Total/NA	Analysis	EPA 537.1		1	441016	WR4P	ELLE	11/08/23 14:27

#### Laboratory References:

ELLE = Eurofins Lancaster Laboratories Environment Testing, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

## Accreditation/Certification Summary

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

### Laboratory: Eurofins Lancaster Laboratories Environment Testing, LLC

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
New York	NELAP	10670	04-01-24

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
537 (Mod)	SPE	Water	6:2 Fluorotelomer sulfonic acid
537 (Mod)	SPE	Water	8:2 Fluorotelomer sulfonic acid
537 (Mod)	SPE	Water	Perfluorobutanoic acid
537 (Mod)	SPE	Water	Perfluorodecanesulfonic acid
537 (Mod)	SPE	Water	Perfluoroheptanesulfonic acid
537 (Mod)	SPE	Water	Perfluoroctanesulfonamide
537 (Mod)	SPE	Water	Perfluoropentanoic acid

## Method Summary

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1

SDG: HOO

Method	Method Description	Protocol	Laboratory
537 (Mod)	EPA 537 Version 1.1 modified	EPA	ELLE
EPA 537.1	EPA 537.1, Ver 1.0 Nov 2018	EPA	ELLE
537.1 DW Prep	Extraction of Perfluorinated Alkyl Acids	EPA	ELLE

**Protocol References:**

EPA = US Environmental Protection Agency

**Laboratory References:**

ELLE = Eurofins Lancaster Laboratories Environment Testing, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

16

## Sample Summary

Client: CT Male Associates DPC  
Project/Site: Hoosick Falls WTP

Job ID: 410-149823-1  
SDG: HOO

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
410-149823-1	GAC Influent	Water	11/02/23 09:50	11/03/23 09:45
410-149823-2	GAC Midfluent	Water	11/02/23 09:55	11/03/23 09:45
410-149823-3	GAC Effluent	Water	11/02/23 10:00	11/03/23 09:45
410-149823-4	PV-2_25	Water	11/02/23 10:05	11/03/23 09:45
410-149823-5	PV-2_50	Water	11/02/23 10:07	11/03/23 09:45
410-149823-6	PV-2_75	Water	11/02/23 10:10	11/03/23 09:45
410-149823-7	FTB01-231102	Water	11/02/23 10:15	11/03/23 09:45
410-149823-8	LTB01-231102	Water	11/02/23 00:00	11/03/23 09:45



Environmental

## Chain of Custody Record

 eurofins  
Environment Testing

410-149823 Chain of Custody

		Sampler: <i>C. Ormsby</i>	Lab FM: Tessier, Kelly	Carrier Tracking No(s):	COC No: 410-77613-21525.1									
Client Contact: Jonathan Dippert, Kirk Moline		Phone:	E-Mail: kelly.tessier@et.eurofinsus.com	State of Origin: <i>NY</i>	Page: <i>50</i> / <i>52</i> Job #:									
Company: CT Male Associates DPC		PWSID:	Analysis Requested											
Address: 50 Century Hill Dr		Due Date Requested:												
City: Latham		TAT Requested (days): <i>Standard</i>												
State, Zip: NY, 12110		Compliance Project: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No												
Phone:		PO #:	Purchase Order not required <i>144756</i>											
Email: j.dippert@ctmale.com, K.Moline@ctmale.com		WO #:												
Project Name: Hoosick Falls WTP		Project #:	41000511											
Site:		SSOW#:												
Sample Identification		Sample Date	Sample Time	Sample Type (C=comp, G=grab) <small>B=Tissue, A=Air</small>	Matrix (W=water, S=solid, O=waste/oil, T=tissue, A=air)	Field Filtered Sample (Yes or No)	Permit Method (Yes or No)	PFC_IDA + (MOD) 7 PFAS Compounds	537_DW - 14 PFAS Drinking Water List	537_DW - 14 PFAS Drinking Water List	Total Number of Containers	Special Instructions/Note:		
<i>GAC INFLUENT</i>		<i>11/2/23</i>	<i>0950</i>	<i>G</i>	<i>Water</i>	<i>N</i>	<i>Y</i>	<i>N</i>			<i>8</i>	<i>PFAS batch QC collected here</i>		
<i>GAC MIDFLUENT</i>			<i>0955</i>		<i>Water</i>		<i>X</i>	<i>X</i>			<i>4</i>			
<i>GAC EFFLUENT</i>			<i>1000</i>		<i>Water</i>		<i>X</i>	<i>X</i>			<i>4</i>			
<i>PV-2-25</i>			<i>1005</i>		<i>Water</i>		<i>X</i>	<i>X</i>			<i>4</i>			
<i>PV-2-50</i>			<i>1007</i>		<i>Water</i>		<i>X</i>	<i>X</i>			<i>4</i>			
<i>PV-2-75</i>			<i>1010</i>		<i>Water</i>		<i>X</i>	<i>X</i>			<i>4</i>			
<i>FTB 01-231102</i>			<i>1015</i>		<i>Water</i>		<i>X</i>	<i>X</i>			<i>4</i>			
<i>LTB 01-231102</i>			<i>-</i>		<i>Water</i>		<i>X</i>	<i>X</i>			<i>4</i>			
					<i>Water</i>									
					<i>Water</i>									
					<i>Water</i>									
					<i>Water</i>									
Possible Hazard Identification						Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)								
<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological						<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For Months								
Deliverable Requested: I, II, III, IV, Other (specify) <i>ASP-B EQUIC 1 file</i>						Special Instructions/QC Requirements:								
Empty Kit Relinquished by:		Date:	Time:	Method of Shipment:										
<i>Christina Ormsby</i>		<i>11/2/23 1530</i>												
Relinquished by		Date/Time:	Company:	Received by			Date/Time:	Company						
Relinquished by		Date/Time:	Company:	Received by			<i>MNR</i>	Date/Time: <i>11/3/23 0945</i>	Company <i>MNR</i>					
Custody Seals Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No.:			Cooler Temperature(s) °C and Other Remarks:			<i>Raw 0.9-10.0.9</i>						

mb

## Login Sample Receipt Checklist

Client: CT Male Associates DPC

Job Number: 410-149823-1

SDG Number: HOO

**Login Number:** 149823

**List Source:** Eurofins Lancaster Laboratories Environment Testing, LLC

**List Number:** 1

**Creator:** Ballard, Megan

Question	Answer	Comment	
The cooler's custody seal is intact.	True		1
The cooler or samples do not appear to have been compromised or tampered with.	True		2
Samples were received on ice.	True		3
Cooler Temperature acceptable,where thermal pres is required(</=6C, not frozen).	True		4
Cooler Temperature is recorded.	True		5
WV:Container Temp acceptable,where thermal pres is required (</=6C, not frozen).	N/A		6
WV: Container Temperature is recorded.	N/A		7
COC is present.	True		8
COC is filled out in ink and legible.	True		9
COC is filled out with all pertinent information.	True		10
There are no discrepancies between the containers received and the COC.	True		11
Sample containers have legible labels.	True		12
Containers are not broken or leaking.	True		13
Sample collection date/times are provided.	True		14
Appropriate sample containers are used.	True		15
Sample bottles are completely filled.	True		16
There is sufficient vol. for all requested analyses.	True		
Is the Field Sampler's name present on COC?	True		
Sample custody seals are intact.	True		
VOA sample vials do not have headspace >6mm in diameter (none, if from WV)?	N/A		

### Sample Preservation Checks (performed by the laboratory)

Question	Answer	Comment	
Did the sample containers checked meet expected preservation conditions?	False	Refer to Job Narrative for details.	